

EXHIBIT 1

**To Plaintiffs' Memorandum Re
Relevancy and Discoverability of FDA
Inspection and Warning Letter and
Recovery Cone Removal System**



DEPARTMENT OF HEALTH AND HUMAN SERVICES

Food and Drug Administration
Los Angeles District
Pacific Region
19701 Fairchild
Irvine, CA 92612

Telephone: 949-608-2900
Fax: 949-608-4415

WARNING LETTER

VIA UNITED PARCEL SERVICE
SIGNATURE REQUIRED

July 13, 2015

W/L # 27-15

Timothy M. Ring
Chairman and Chief Executive Officer
C.R. Bard Inc.
730 Central Ave.
Murray Hill, NJ 07974

Dear Mr. Ring:

During inspection of your C.R. Bard Inc. facility located at 289 Bay Rd, Queensbury, NY, on October 6, 2014, through November 25, 2014, and during inspection of your Bard Peripheral Vascular facility located at 1625 W. 3rd St., Tempe, AZ, on November 18, 2014, through January 05, 2015, investigators from the United States Food and Drug Administration (FDA) determined that your firm is a specification developer and manufacturer for the Inferior Vena Cava (IVC) filter delivery systems and components, including, but not limited to, the Denali Filter, the Simon Nitinol Filter and Recovery Cone Removal Kit. This warning letter addresses violations found at the Bard Peripheral Vascular facility located at 1625 W. 3rd St., Tempe, AZ and C.R. Bard Inc. facility located at 289 Bay Rd, Queensbury, NY. Under section 201(h) of the Federal Food, Drug, and Cosmetic Act (the Act), 21 U.S.C. § 321(h), these products are devices because they are intended for use in the diagnosis of disease or other conditions or in the cure, mitigation, treatment, or prevention of disease, or is intended to affect the structure or any function of the body.

We received responses dated December 17, 2014, January 15, 2015, February 18, 2015, March 16, 2015, April 17, 2015, and May 6, 2015, from Mr. Jason J. Gaede, Vice President Plant Operations, C.R. Bard Inc., Queensbury, NY. We also received responses dated January 26, 2015, February 26, 2015, March 26, 2015, April 24, 2015, and May 22, 2015, from Steve S. Williamson, President, Bard Peripheral Vascular, a Division of C.R. Bard, Tempe, AZ. These were responses to the observations noted on Form FDA 483s, Lists of Inspectional Observations that were issued to you at the close of our inspections. We address your responses below, in relation to each of the noted violations. These violations include, but are not limited to, the following:



BPV-17-01-00204231

C.R. Bard Inc.
Page 2

Adulteration/Misbranding Violations at the Tempe, AZ facility

1. FDA has learned that your firm manufactures the Recovery Cone Removal System, Model RC-15 in the United States without marketing clearance or approval, in violation of the Act. Under section 201(h) of the Act, 21 U.S.C. § 321(h), this product is a device because it is intended for use in the diagnosis of disease or other conditions or in the cure, mitigation, treatment, or prevention of disease, or is intended to affect the structure or any function of the body. As explained below, this device is being marketed without the required clearance or approval.

The Recovery Cone Removal System, Model RC-15 is adulterated under section 501(f)(1)(B) of the Act, 21 U.S.C. § 351(f)(1)(B), because you do not have an approved application for premarket approval (PMA) in effect pursuant to section 515(a) of the Act, 21 U.S.C. § 360e(a), or an approved application for an investigational device exemption (IDE) under section 520(g) of the Act, 21 U.S.C. § 360j(g). The Recovery Cone Removal System, Model RC-15 is also misbranded under section 502(o) the Act, 21 U.S.C. § 352(o), because you did not notify the agency of your intent to introduce the device into commercial distribution, as required by section 510(k) of the Act, 21 U.S.C. § 360(k).

FDA reviewed labeling for the Recovery Cone Removal System, Model RC-15, which revealed that this device is intended to percutaneously remove the Recovery Filter, Recovery G2 Filter and the G2X Filter as indicated. FDA is aware that your firm submitted both in-vitro and in-vivo testing demonstrating the use of the Recovery Cone Removal System, Model RC-15 for removal of the Recovery Filter (K031328), the G2X Filter (K082305), the G2 Express Filter (K080668), and the G2 Filter (K073090). However, the Recovery Cone System, Model RC-15 was not included as part of the clearances for any of the aforementioned IVC filters. Therefore, your firm is marketing the Recovery Cone Removal System, Model RC-15 in the United States without marketing clearance or approval. Percutaneous retrieval systems, such as the Recovery Cone Removal System, Model RC-15, are regulated as manual surgical instruments intended for specialized use within a specific medical specialty, and thus require marketing authorization in order to be legally marketed in the United States.

Your firm has not submitted any correspondence to FDA regarding this violation to date.

2. FDA has also learned that your firm manufactures the Recovery Cone Removal System, Model FBRC in the United States without marketing clearance or approval, in violation of the Act. Under section 201(h) of the Act, 21 U.S.C. § 321(h), this product is a device because it is intended for use in the diagnosis of disease or other conditions or in the cure, mitigation, treatment, or prevention of disease, or intended to affect the structure or any function of the body. As explained below, this device is being marketed without the required clearance or approval.

The Recovery Cone Removal System, Model FBRC is adulterated under section 501(f)(1)(B) of the Act, 21 U.S.C. § 351(f)(1)(B), because you do not have an approved

C.R. Bard Inc.

Page 3

application for premarket approval (PMA) in effect pursuant to section 515(a) of the Act, 21 U.S.C. § 360e(a), or an approved application for an investigational device exemption (IDE) under section 520(g) of the Act, 21 U.S.C. § 360j(g) for the device as described and marketed. The Recovery Cone Removal System, Model FBRC is also misbranded under section 502(o) of the Act, 21 U.S.C. 352(o), because you introduced or delivered into interstate commerce for commercial distribution a device with major changes/modifications to the intended use without submitting a new premarket notification to the agency as required by section 510(k), 21 U.S.C. 360(k), and 21 C.F.R. 807.81(a)(3)(ii).

You have listed the Recovery Cone Removal System, Model FBRC as a class I surgical snare under 21 CFR 878.4800. Devices classified under 21 CFR 878.4800 (Surgical Instrument, Manual) are exempt from premarket notification, unless they exceed the limitations on exemptions at 21 CFR 878.9(a). However, there is evidence that the Recovery Cone Removal System, Model FBRC is intended for uses that are different from those of legally marketed devices under 21 CFR 878.4800 (Surgical Instrument, Manual). Devices of this type usually consist of a non-powered, hand-held, or hand-manipulated device that is either reusable or disposable, which are intended to be used in general surgical procedures. Manual surgical instruments intended for specialized uses within a specific medical specialty are classified under regulations separate from 21 CFR 878.4800, depending on the labeled specialized use of the device. However, your firm is marketing the Recovery Cone Removal System, Model FBRC for a specialized intended use, namely percutaneous removal of inferior vena cava filters, specifically your firm's G2X Filter, G2 Express Filter, and G2 Filter. The labeling for the Recovery Cone Removal System, Model FBRC also indicates that your product is intended to percutaneously remove a foreign body.

Based on the above, FDA believes that the Recovery Cone Removal System, Model FBRC is regulated as a percutaneous retrieval system, which is a manual surgical instrument intended for specialized use within a specific medical specialty, cardiovascular surgery. Because there is evidence that the Recovery Cone Removal System, Model FBRC is intended for uses that are different from those of legally marketed devices classified under 21 CFR 878.4800, it exceeds the limitations described in 21 C.F.R. 878.9(a) and is not exempt from premarket notification.

Your firm has not submitted any correspondence to the FDA regarding this violation to date. For a device requiring premarket approval, the notification required by section 510(k) of the Act, 21 U.S.C. § 360(k), is deemed satisfied when a PMA is pending before the agency. 21 C.F.R. 807.81(b). The kind of information you need to submit in order to obtain approval or clearance for your device is described on the Internet at <http://www.fda.gov/cdrh/devadvice/3122.html>. The FDA will evaluate the information you submit and decide whether your product may be legally marketed.

C.R. Bard Inc.
Page 4

FDA requests that Bard Peripheral Vascular immediately cease activities that result in the misbranding or adulteration of the Recovery Cone Removal System, Model RC-15 and the Recovery Cone Removal System, Model FBRC, such as the commercial distribution of the devices for the uses discussed above.

Quality System Violations

The inspections also revealed that these devices are adulterated within the meaning of section 501(h) of the Act, 21 U.S.C. § 351(h), in that the methods used in, or the facilities or controls used for, their manufacture, packing, storage, or installation are not in conformity with the current good manufacturing practice requirements of the Quality System regulation found at Title 21, Code of Federal Regulations (CFR), Part 820.

The inspection of your Bard Peripheral Vascular Facility located at 1625 W. 3rd St., Tempe, AZ also revealed that the IVC Denali Filter Delivery System is misbranded under Section 502(t)(2) of the Act, 21 U.S.C. § 352(t)(2), in that your firm failed or refused to furnish material or information regarding the devices that is required by or under Section 519 of the Act, 21 U.S.C. § 360i, and 21 CFR Part 803 - Medical Device Reporting.

Quality System Regulation Violations at the Tempe, AZ facility and Queensbury, NY facility

3. Failure to establish and maintain procedures for receiving, reviewing, and evaluating complaints as required by 21 CFR 820.198(a). Your current procedures governing complaint investigation activities at your facilities, CQA-STD-24 Standard for Product Complaint Handling Rev. 11 and CQA-STD-55, Standard for Complaint Investigation Process Rev. 01, MP9113, Complaint Investigation Activity Rev. 40, SOPQ0153100, BPV Complaint Handling System, Rev. 18, and SOPQ0700200, Complaint Investigation Procedures, Rev. 15 do not ensure product complaints are adequately evaluated. For example:
 - a. Your current procedures governing complaint investigation activities, CQA-STD-24 Standard for Product Complaint Handling Rev. 11, CQA-STD-55, Standard for Complaint Investigation Process Rev. 01, MP9113, Complaint Investigation Activity Rev. 40, SOPQ0153100, BPV Complaint Handling System, Rev. 18, and SOPQ0700200, Complaint Investigation Procedures, Rev. 15 do not include adequate instructions for ensuring that complaints involving a device or device component provided by a supplier are adequately evaluated for root cause of the alleged device failure and that appropriate corrective action is implemented with your suppliers.
 - b. Complaint 562535 for a G2 Filter, embolization of a detached filter arm with associated areas of hemorrhage and necrosis in the right lung was filed as a malfunction Medical Device Report [MDR] and should have been filed as a death. The following complaints were filed as malfunctions and should have been filed as serious injuries: Complaint 628924, Eclipse Filter, detached filter limb resulting in pericardial effusion and cardiac catheterization; 574429, G2 Express

C.R. Bard Inc.
Page 5

Filter, broken filter and surgical intervention; 602904, Denali Jugular System, detached filter arm embedded in IVC wall with filter retrieval; 443237, G2 Filter, detached filter limb in renal vein with IVC wall perforation and blood thinner treatment; 446809, G2 Express Filter, IVC perforation and aneurysm; 454485, G2 Filter, abdominal pain with filter legs protruding through IVC wall and percutaneous removal; 562566, G2 Filter, abdominal pain with filter legs perforating IVC wall, partial retrieval and residual filter leg fragment embedded in IVC wall.

- c. Complaints 507112, 507109, 507115, 507252, 507280, 507302, 507311 and 507325 report at least 10 patients who were exposed to scheduled retrieval surgical procedures to remove an IVC filter that were not successful. However, these complaint files do not document sufficient information to allow for adequate complaint investigation and disposition, including, but not limited to, MDR determination. For example, the complaints do not include information regarding prolonged or repeated surgery that may have occurred as a result of failed attempts to remove the filters, information regarding why the filters were scheduled to be removed and potential complications related to leaving them in the patient due to failed removal, and/or if any additional drugs or anesthetics had to be supplied to the patients.

We find that your responses dated December 17, 2014, January 15, 2015, February 18, 2015, March 16, 2015, April 17, 2015 and May 6, 2015 from Mr. Jason J. Gaede, Vice President Plant Operations, C.R. Bard Inc., Queensbury, NY and your responses dated January 26, 2015, February 26, 2015, March 26, 2015, April 24, 2015, and May 22, 2015, from Mr. Steve S. Williamson, President, Bard Peripheral Vascular, a Division of C.R. Bard, Tempe, AZ do not adequately address these deficiencies. For example, your January 26, 2015, response states that you made clerical errors and that you opened a CAPA to track training and determination of root cause with corrective and preventive actions. Your response is inadequate and does not assure that your complaint handling system reviews and evaluates complaints adequately. Additionally, the revised complaint procedures provided with your initial responses do not include adequate corrections to complaint investigation procedures with regards to the above stated deficiencies. Your follow-up responses do not address any corrections for complaint handling deficiencies. Your responses also state that all actions have been implemented with respect to the violation and that your firm considers your response to be complete.

Quality System Regulation Violations at the Queensbury, NY facility

4. Failure to validate, with a high degree of assurance and approve according to established procedures, a manufacturing process that cannot be fully verified by subsequent inspection and testing, to ensure the process will continue to meet specifications as required by 21 CFR 820.75(a).
 - a. Specifically, IVC filter cleaning, to include removal of chemical processing contaminants, has not been validated for IVC Filters to include Simon Nitinol Filters, Eclipse Filters and Denali Filters. For example, production of Denali

C.R. Bard Inc.
Page 6

Filters requires the use of several processing agents, including, but not limited to the following: nitric acid, methanol, sulfamic acid solution, thermo quench salt, glycolic acid, citric acid, and/or hydrofluoric acid. The cleaning processes for IVC filters are not validated or otherwise verified to demonstrate that the above substances are reduced to acceptable levels during routine processing under worst case conditions. Therefore, your manufacturing process was not validated with a high degree of assurance and approved according to established procedures, nor were the process results fully verified by subsequent inspection and test, as 21 CFR 820.75(a) requires.

- b. Your firm's own Process Qualification (PQ) Final Report PRT10-064 Rev A., dated May 29, 2013, states that a 100% inspection plan is necessary for all failed predefined acceptance criteria during process validation. In particular, the PQ Final Report defines process capability acceptance criteria for Denali Filter Part #RM5109000 dimensions of C, D, L, M, G, N, W, F, T, U and the radial force functional test to be a CpK greater than or equal to 1.33 as a requirement to validate the process. Your process qualification failed to meet this predefined acceptance criteria for these filter dimensions and functional test. Therefore, according to your own manufacturing process validation document, 100% inspection for verification of these specifications on each lot of product is required to mitigate your failed process validation. However, your firm has not ensured 100% inspection of dimensions N, W, F, G and M, which lacked validation. Your firm has also not conducted adequate subsequent process validation studies to eliminate this requirement. As a result, your manufacturing process was not validated with a high degree of assurance and approved according to established procedures, nor were the process results fully verified by subsequent inspection and test, as 21 CFR 820.75(a) requires.

We find that your responses dated December 17, 2014, January 15, 2015, February 18, 2015, March 16, 2015, April 17, 2015, and May 6, 2015, are not adequate for the following reasons:

- With regard to your promised corrections for IVC filter cleaning lacking process validation, we find your response partially adequate. We acknowledge your firm's actions to date associated with the CAPA you opened in response to this observation, GF-CAPA-15-002. We acknowledge that you reviewed 510K submission data for the Denali Filters, conducted recent cytotoxicity testing for Denali Filters and revised your Process Validation procedure, SOP 820.100.6, Rev 43., to specifically include the requirement of validating cleaning processes for components or devices that undergo contact with processing agents. We acknowledge your progress to date validating the cleaning processes for Denali Filters manufactured by both of your suppliers and Simon Nitinol Filters; however, this data will need further review during a follow-up inspection to verify adequacy of actions taken. We also acknowledge your performance of exhaustive extraction testing for the Denali Filters manufactured by one of your suppliers; however, the other supplier of these filters uses a different manufacturing process, processing agents, and equipment. Because of these differences, we recommend that you perform exhaustive extraction testing for Denali Filters manufactured by this supplier to ensure no residuals are present on these devices. Additionally, your firm has stated it is no longer manufacturing the Eclipse Filters as of 9/8/14; however, your firm has not indicated plans

C.R. Bard Inc.
Page 7

regarding the stored inventory of these devices and continues to market them in the United States. Your response to date does not indicate corrective action for these devices that may still be in inventory and/or may still be distributed.

- With regard to your promised corrections relating to process validation of your Denali Filter Part #RM5109000 dimensions of F, N, W, G and M, we find that your response is not adequate. Your Process Qualification (PQ) Final Report PRT10-064 Rev A., dated May 29, 2013, states that a 100% inspection plan is necessary for all failed predefined acceptance criteria during process validation. In particular, the PQ Final Report defines process capability acceptance criteria for Denali Filter Part #RM5109000 dimensions and/or functional tests of C, D, L, M, G, N, W, F, T, U and Radial Force to be a CpK greater than or equal to 1.33 as a requirement to validate the process. This predefined acceptance criterion was not met. Consequently, your firm conducted a retrospective analysis to change this original process validation criterion. You state your firm retrospectively analyzed data and determined that dimensions N and W may stay on AQL 0.65 limited inspections because the analysis demonstrated a 95/99.9% confidence level. However, your rationale for changing the original process validation acceptance criterion for dimensions N and W is not adequately supported.

Further, you have not been successful at validating the manufacturing process with respect to dimensions F, G and M, which failed predefined acceptance criteria. Your firm has not provided adequate data to support that these dimensions have been 100% inspected for every lot of product manufactured, as required by your firm's own manufacturing process validation document, PQ Final Report PRT10-064. And lastly, the corrective actions proposed as part of CAPA GF-CAPA-15-001 are in progress, and will need verification of implementation upon completion during a future inspection. For these reasons, your responses dated December 17, 2014, January 15, 2015, February 18, 2015, March 16, 2015, April 17, 2015, and May 6, 2015, are inadequate.

5. Failure to establish and maintain procedures for acceptance of incoming product and to inspect, test or otherwise verify incoming product as conforming to specified requirements as required by 21 CFR 820.80(b).

Specifically, based on your Process Qualification (PQ) Final Report PRT10-064 Rev A., dated May 29, 2013, process capability acceptance criteria of CpK greater than or equal to 1.33 for Denali Filter Part #RM5109000 dimensions C, D, L, M, G, N, W, F, T, U and the radial force functional test were not met. As a result, these dimensions and/or functional tests were to remain on a 100% inspection plan during manufacture at your supplier in order to be accepted into inventory. However, your firm accepted supplier lot numbers #33511-39, #33511-42, #33511-43, #33511-44, #33511-45, #33511-46, #33511-47, #33511-48, #33511-49, #33511-50, #33511-51, #33511-53, #33595-2A-20, #33595-2A-27, #33595-2A-29, #33595-2A-31, #33595-2A-30, #33595-2A-33, #33595-2A-35, #33595-2A-39, #33595-2A-40, #33595-4, and #33771-112 of Denali Filter Part #RM5109000, which your supplier inspected with an AQL 0.65 sampling plan for dimensions N, W, F, G, and M, rather than 100% inspection. Your firm also accepted

C.R. Bard Inc.
Page 8

supplier lot numbers #33771-129 and #34360-24 of Denali Filter Part #RM5109000, inspected by your supplier with an AQL 0.65 sampling plan for dimensions N and W, rather than 100% inspection. Your procedures for acceptance of incoming product, including Inspection Plan IP #1450935 Rev. 2., were not adequately established and maintained to verify that incoming product conformed to your specified requirements of 100% inspection plan for dimensions N, W, F, G, and M. As a result, you failed to inspect, test or otherwise verify incoming product as conforming to your specified acceptance requirements, as required by 21 CFR 820.80(b).

We find your responses dated December 17, 2014, January 15, 2015, February 18, 2015, March 16, 2015, April 17, 2015, and May 6, 2015, partially adequate. Your responses do not clarify whether acceptable corrective actions have been taken with the above stated lots of Denali Filter components that lacked 100% inspection of dimensions N, W, F, G, and M to ensure your specified acceptance requirements have been met for these accepted lots. Your response does not contain evidence that the above stated lots indicating an AQL 0.65 sampling plan for dimensions F, G and M were in fact inspected at 100% for F, G and M. Further, your response does not contain evidence that your supplier's AQL 0.65 sampling plan is an adequate inspection, test, or verification of incoming product for dimensions N and W. We acknowledge your firm has opened CAPA GF-CAPA-15-005 to address systemic corrections to this observation; however, outputs of this CAPA are still in progress and will need to be verified during an FDA inspection of your firm.

6. Failure to establish and maintain procedures to ensure that all purchased or otherwise received product and services conform to specified requirements, as required by 21 CFR 820.50. In particular, 21 CFR 820.50(a) requires that each manufacturer establish and maintain requirements, including quality requirements, that must be met by suppliers, contractors, and consultants.

Specifically, your Process Qualification (PQ) Final Report PRT10-064 Rev A., dated May 29, 2013, documented that process capabilities for filter dimensions C, D, L, M, G, N, W, F, T, U and the functional test radial force, as defined under section 7 Acceptance Criteria – Process Qualification of this report, were not met for all dimensions and functional tests. The PQ Final Report states that because the process capabilities were not met, these filter dimensions and functional test should remain on a 100% inspection plan at your supplier until such time that objective evidence indicates process capability has been demonstrated. However, your supplier failed to inspect the product on a 100% inspection plan for filter dimensions N, W, F, G and M, and process capability was not demonstrated through objective evidence. For example, from approximately May 11, 2013 to August 5, 2013, your supplier of Denali Filter Part #RM5109000 provided you with a Certificate of Compliance for supplier lot numbers #33511-39, #33511-42, #33511-43, #33511-44, #33511-45, #33511-46, #33511-47, #33511-48, #33511-49, #33511-50, #33511-51, #33511-53, #33595-2A-20, #33595-2A-27, #33595-2A-29, #33595-2A-31, #33595-2A-30, #33595-2A-33, #33595-2A-35, #33595-2A-39, #33595-

C.R. Bard Inc.
Page 9

2A-40, #33595-4, and #33771-112. These certificates documented the supplier did not conduct 100% inspection for filter dimensions N, W, F, G and M. Your firm did not begin to address the issue with your supplier until approximately August 5, 2013 during an audit of your supplier, which was after most of these 23 lots were accepted into your inventory and used in the manufacture of finished Denali IVC Filter devices.

In the PQ Final Report PRT10-064 Rev A., your firm establishes purchasing control procedures as required by 21 CFR 820.50(a). These procedures include continued 100% inspection by your supplier for process capabilities that were not met with regards to filter dimensions C, D, L, M, G, N, W, F, T, U and for process capabilities that were not met with respect to the functional test radial force. However, when filter dimensions N, W, F, G, and M failed your established process capabilities, your supplier did not conduct 100% inspection. By failing to maintain adequate supplier control procedures (i.e., by failing to ensure 100% inspection was conducted for failed process capabilities), your firm violated 21 CFR 820.50(a), which requires that manufacturers establish and maintain requirements that must be met by suppliers.

Additionally, when suppliers are placed on Limited Approved status, such as your supplier of the Denali Filter Part #RM5109000, Norman Noble Inc., you do not have adequate instructions in your supplier control procedures, including but not limited to CQA-STA-18 Supplier Performance Management Rev. 05, to re-evaluate suppliers to ensure that the supplier is better able to meet your specifications.

We find that your responses December 17, 2014, January 15, 2015, February 18, 2015, March 16, 2015, April 17, 2015, and May 6, 2015, appear adequate, but are still in progress and will need to be verified during an FDA inspection of your firm.

MDR Violations at the Tempe, AZ facility

Our inspection of your Bard Peripheral Vascular facility located at 1625 W. 3rd St., Tempe, AZ also revealed that the Cardiovascular intravascular filter, (IVC Denali Filter Delivery System), is misbranded under Section 502(t)(2) of the Act, 21 U.S.C. § 352(t)(2), in that your firm failed or refused to furnish material or information regarding the devices that is required by or under Section 519 of the Act, 21 U.S.C. § 360i, and 21 CFR Part 803 - Medical Device Reporting. Significant deviations include, but are not limited to:

7. Failure to submit a report to FDA no later than 30 calendar days after the day that your firm received or otherwise became aware of information, from any source, that reasonably suggests that a device that your firm markets has malfunctioned and this device or a similar device that your firm markets would be likely to cause or contribute to a death or serious injury, if the malfunction were to recur, as required by 21 CFR 803.50(a)(2).

C.R. Bard Inc.
Page 10

For example, Complaint numbers 574292, 427438, 418704, and 416808 describe a malfunction of your firm's device, which is classified as a long term implant. Your firm did not rule out that the reported malfunctions would not be likely to cause or contribute to a death or serious injury, if it were to recur. Therefore, an MDR should have been submitted for each of the referenced complaints.

We reviewed your firm's responses received by the FDA, including the January 26, 2015, response, and conclude the response is not adequate. Your firm did not submit MDRs for the above referenced complaints and failed to justify why such malfunctions would not be likely to cause or contribute to a death or serious injury, if the malfunctions were to recur.

8. Failure to obtain and submit to the FDA information that is incomplete or missing from reports submitted by user facilities, importers, and other initial reporters; and if unable to submit complete information on a report, failure to provide a statement in your firm's report explaining why required information was incomplete and the steps taken by your firm to obtain the information, as required by 21 CFR 803.50(b)(2) and 21 CFR 803.50(b)(3).

Specifically, your firm submitted 15 MDRs to the FDA, which did not identify the patient's "Age at Time of Event" or "Date of Birth" in Blocks A2 and A4, respectively, of the FDA Form 3500A. In addition, your firm did not include an explanation of why the required information was not provided and the steps taken to obtain such information.

We reviewed your firm's responses received by the FDA, including the January 26, 2015, response, and conclude the response is not adequate. Although the FDA has received supplement reports for some of the MDRs, we have not received supplements for all.

The eMDR Final Rule requiring manufacturers and importers to submit electronic Medical Device Reports (eMDRs) to FDA was published on February 13, 2014. The requirements of this final rule will take effect on August 14, 2015. If your firm is not currently submitting reports electronically, we encourage you to visit the following web link for additional information about the electronic reporting requirements:

<http://www.fda.gov/ForIndustry/FDAeSubmitter/ucm107903.htm>

If your firm wishes to discuss MDR Reportability criteria or to schedule further communications, it may contact the Reportability Review Team by email at:

ReportabilityReviewTeam@fda.hhs.gov

Your firm should take prompt action to correct the violations addressed in this letter. Failure to promptly correct these violations may result in regulatory action being initiated by the FDA without further notice. These actions include, but are not limited to, seizure, injunction, and civil money penalties. Also, federal agencies may be advised of the issuance of Warning Letters about devices so that they may take this information into account when considering the award of contracts. Additionally, premarket approval applications for Class III devices to which the Quality

C.R. Bard Inc.
Page 11

System regulation violations are reasonably related will not be approved until the violations have been corrected. Requests for Certificates to Foreign Governments will not be granted until the violations related to the subject devices have been corrected.

Please notify this office in writing within fifteen (15) business days from the date you receive this letter of the specific steps your firm has taken to correct the noted violations, as well as an explanation of how your firm plans to prevent these violations, or similar violations, from occurring again. Include documentation of the corrections and/or corrective actions (including any systemic corrective actions) that your firm has taken. If your firm's planned corrections and/or corrective actions will occur over time, please include a timetable for implementation of those activities. If corrections and/or corrective actions cannot be completed within fifteen (15) business days, state the reason for the delay and the time within which these activities will be completed. Your firm's response should be comprehensive and address all violations included in this Warning Letter.

Your written response should be sent to the Food and Drug Administration; Attention:

Dr. Raymond W. Brullo
Compliance Officer, Los Angeles District
U. S. Food and Drug Administration
19701 Fairchild
Irvine, CA 92612

A copy of your written response should also be sent to:


LCDR Catherine M. Beer
Compliance Officer
U. S. Food and Drug Administration
One Winners Circle, Suite 110
Albany, NY 12205

If you have any questions about the content of this letter please contact: Dr. Raymond W. Brullo at (949) 608-2918.

Finally, you should know that this letter is not intended to be an all-inclusive list of the violations at your facility. It is your responsibility to ensure compliance with applicable laws and regulations administered by FDA. The specific violations noted in this letter and in the Form FDA 483, Inspectional Observations (FDA 483), issued at the close out of the inspection may be symptomatic of serious problems in your firm's manufacturing and quality management systems. You should investigate and determine the causes of the violations, and take prompt actions to correct the violations and to bring your products into compliance.

C.R. Bard Inc.
Page 12

Sincerely yours,


Alonza E. Cruse, Director
Los Angeles District

Cc:

Kevin J. Bovee
Director of Quality Assurance
C.R. Bard, Inc.
289 Bay Road
Queensbury, NY 12804

Jason J. Gaede
Vice President, Plant Operations
C.R. Bard, Inc.
289 Bay Road
Queensbury, NY 12804

Mark M. Walaska
Staff Vice President Manufacturing
Bard Peripheral Vascular, Inc.
1625 W. 3rd St.
Tempe, AZ 85281

Steve S. Williamson
President
Bard Peripheral Vascular, Inc.
1625 W. 3rd St.
Tempe, AZ 85281

Patricia Christian
Vice President, Quality, Regulatory and Medical Affairs
C.R. Bard, Inc.
730 Central Ave.
Murray Hill, NJ 07974

Gin Schulz
Vice President, Corporate Quality Assurance
C.R. Bard, Inc.
730 Central Ave.
Murray Hill, NJ 07974

BPV-17-01-00204241

C.R. Bard Inc.
Page 13

bcc:

NYK-DO CB: (Beer/Alexander)
NYK-DO IB: (Frankovic/Terzian/Warner/Izyk)
CFSAN via CMS
HFI-35 (purged) via CMS
Legal file
CMS Case # 453127
EI file C.R. Bard, Queensbury NY (1313046)

LOS-DO DCB file
LOS-DO IB [M. Maxwell, S. Perkins, M. Saale – e-copies]
CMS Case # 453127 [issued version for CO redaction]
Legal File
Factory File
Chron [Brullo]

EXHIBIT 2

**To Plaintiffs' Memorandum Re
Relevancy and Discoverability of FDA
Inspection and Warning Letter and
Recovery Cone Removal System**

1 IN THE UNITED STATES DISTRICT COURT

2 FOR THE DISTRICT OF ARIZONA

3

4

IN RE: BARD IVC FILTERS PRODUCTS)

5 LIABILITY LITIGATION) MD No.: 02641

_____)

6

7

8

9

10 DO NOT DISCLOSE - SUBJECT TO FURTHER CONFIDENTIALITY REVIEW

11

12

13 CONTINUED VIDEOTAPED DEPOSITION OF CHAD MICHAEL MODRA

14

15

Phoenix, Arizona

16 January 20, 2016

9:01 a.m.

17

18

19

20

21

22

23 REPORTED BY:

24 Robin L. B. Osterode, RPR, CSR

25 AZ Certified Reporter No. 50695

1 (Previously marked Exhibit 102 is
2 attached hereto.)

3 BY MR. LOPEZ:

4 Q. Exhibit 102 was marked --

5 A. Yeah.

6 Q. Do you have it?

7 A. Yes, I do.

8 Q. So we've been talking about this warning
9 letter, and that's what this deposition is about.
10 This is the warning letter, July 13, 2015. Correct?

11 A. Correct.

12 Q. And I think last time we agreed that the
13 warning letter is really a letter that involves
14 multiple violations of safety processes, citations,
15 because it actually says that in the warning letter.
16 Right?

17 A. It does say that.

18 Q. So why do you suppose they call it a
19 warning letter?

20 A. They wanted to send a message to the firm
21 that their -- you know, they take these certain
22 elements seriously or they're concerned about these
23 elements of a 483 item form. So that's their vehicle
24 of communicating that to us.

25 Q. Well, they actually -- I think they include

EXHIBIT 3

**To Plaintiffs' Memorandum Re
Relevancy and Discoverability of FDA
Inspection and Warning Letter and
Recovery Cone Removal System**

1 IN THE UNITED STATES DISTRICT COURT

2 FOR THE DISTRICT OF ARIZONA

3

4

IN RE: BARD IVC FILTERS PRODUCTS)

5 LIABILITY LITIGATION) MD No.: 02641

_____)

6

7

8

9

10 DO NOT DISCLOSE - SUBJECT TO FURTHER CONFIDENTIALITY REVIEW

11

12

13 CONTINUED VIDEOTAPED DEPOSITION OF CHAD MICHAEL MODRA

14

15

Phoenix, Arizona

16 January 20, 2016

9:01 a.m.

17

18

19

20

21

22

23 REPORTED BY:

24 Robin L. B. Osterode, RPR, CSR

25 AZ Certified Reporter No. 50695

1 the warning at the end of the letter, if you look at
2 Bates number 242.

3 A. Yes.

4 Q. Okay. As a matter of fact, they also make
5 it clear that this letter is not intended to be an
6 all-inclusive list of the violations at your
7 facility?

8 A. Correct.

9 Q. And it's your responsibility, meaning the
10 company, to ensure compliance with applicable laws
11 and regulations administered by FDA. Right?

12 A. Yes.

13 Q. In other words, FDA is saying that, you
14 know, we didn't come through and do a
15 corner-to-corner, you know, inch-by-inch audit and
16 evaluation of your company; we just did certain
17 things, and we found these violations. And these are
18 not to be considered an all-inclusive list of these
19 violations. Right?

20 A. That's correct.


21 Q. And they're telling -- the warning here is
22 that you probably should be looking at other places
23 if you haven't yet, and not rely on FDA to find them?

24 A. Of course.

25 Q. And that you should do that irrespective of

EXHIBIT 4

**To Plaintiffs' Memorandum Re
Relevancy and Discoverability of FDA
Inspection and Warning Letter and
Recovery Cone Removal System**



To: T. Ring/J. Weiland

From: C. Ganser

Date: August 3, 2005

Subject: IVC Recovery Filter Adverse Events (Migrations/Fractures) – Executive Summary

Following is an adverse event summary of the Bard IVC Recovery Filter for migrations and fractures through 8/2/2005:

Migrations

- 49 filter migrations reported > 2 cm
- 28/49 cases confirmed migration of the filter encased in large thrombi
- 12/49 cases - clot presence was unknown
- 15/28 migrations with confirmed thrombi encasement migrated to the heart
- 16/49 migrations - involved morbidly obese/gastric bypass patients
- 13/49 migrations - death reported
- 11/13 migration associated deaths - massive clot burden reported to have overwhelmed the filter (2 had unknown clot burdens)
- 8/13 migration associated deaths - involved morbidly obese/gastric bypass patients
- 1/13 migration associated deaths - involved Australian patient, information pending
- 1/13 migration associated deaths - involved woman suffering from a sub-arachnoid hemorrhage
- 1/13 migration associated deaths - involved trauma patient (massive head wound)
- 1/13 migration associated deaths - involved a patient post Achilles tendon surgery
- 1/13 migration associated deaths - investigation revealed insufficient data
- 5 pulmonary embolism fatalities, unrelated to migration
- 33,178 units sold as of 8/2/2005
- Estimated 26,542 units placed
- Bard IVC Recovery filter migration rate 0.148% (based on units sold)
- Bard IVC Recovery filter migration related mortality rate 0.039% (based on units sold)

Comparative MAUDE/IMS data for IVC filter fatalities (through Q2 2005):

Rates	Fatalities	Migration
SNF	0.0000%	0.0027%
Recovery	0.0563%	0.1220%
Vena Tech	0.0082%	0.0532%
Greenfield	0.0077%	0.0242%
Bird's Nest	0.0142%	0.0283%
Tulip	0.0149%	0.0277%
TrapEase	0.0147%	0.0152%
OptEase	0.0204%	0.0204%

MAUDE Fatalities are associated with reports associated with migration, caval perforation, caval thrombosis, pulmonary embolism, and failed deployment.



Confidential

Page 1 of 2

Fractures

- 68 reports of filter fractures have been reported to date (rate of 0.205%)
- 25/68 cases involved fragments migrated to the heart/lung (0.075%)
- 4/25 fractures - the patient required surgery to remove detached limb(s)
- No other injuries have been reported
- A literature review reveals that filter fracture is a known complication of IVC filters, with reported rates in the range of 0.05% - 10%

EXHIBIT 5

**To Plaintiffs' Memorandum Re
Relevancy and Discoverability of FDA
Inspection and Warning Letter and
Recovery Cone Removal System**

	A	B	C	D	E	F	G	H	I	J	K	L	M	N	O	P	Q	R	S	T
1	FILTER SALES (IMS Q1 '00 to Q4 '04, + Trend Q1 - Q3 '05)																			
2																				
3	G2	2,246 * Actual Sales Through November 7, 2005																		
4	SNF	77,572 * Actual Sales through Q3 2005																		
5	Recovery	34,752 * Actual Sales through Q3 2005																		
6	Vena Tech	51,265 Need IMS Sales Data - currently based on trend																		
7	Greenfield	199,460 Need IMS Sales Data - currently based on trend																		
8	Bird's Nest	7,249 Need IMS Sales Data - currently based on trend																		
9	Tulip	50,638 Need IMS Sales Data - currently based on trend																		
10	TrapEase	192,659 Need IMS Sales Data - currently based on trend																		
11	OptEase	16,905 Need IMS Sales Data - currently based on trend																		
12	TOTAL	630,600																		
13																				
14	MAUDE DATA THROUGH Q3 2005																			
15																				
16	Counts	Fatalities	FltEmb Deaths	Nondeploy AEs	Total Deploy	Total AEs	Migration	Filter Embolization	Migration	Caval Perforation	Caval Thrombosis	Pulmonary Embolism	Filter Fracture	Guidewire Entrapment	Misplaced Deployment	Partial Deployment	Failed Deployment	Tilted Filter	Other	Detachment of Component(s)
17	G2 Actual	0	0	4	16	20	1	0	1	3	0	0	0	0	14	0	2	0	0	0
18	SNF	0	0	26	25	51	1	1	2	7	0	0	0	0	4	10	11	0	0	15
19	Recovery	19	13	118	22	140	27	19	46	13	0	11	45	0	2	12	8	1	2	0
20	Vena Tech	4	3	32	21	53	4	23	27	0	0	1	1	1	0	4	16	0	3	0
21	Greenfield	15	3	85	222	307	14	34	48	10	1	4	9	12	41	130	39	1	12	0
22	Bird's Nest	1	0	21	9	30	1	1	2	11	0	1	7	0	6	0	3	0	0	0
23	Tulip	7	2	39	35	74	6	8	14	10	2	2	0	0	5	17	13	4	6	1
24	TrapEase	29	9	147	33	180	10	19	29	19	62	8	13	0	0	15	18	0	11	0
25	OptEase	3	0	23	5	28	4	0	4	0	4	4	8	0	2	0	3	0	3	0
26																				
27	Rates	Fatalities	FltEmb Deaths	Nondeploy AEs	Total Deploy	Total AEs	Migration	Filter Embolization	Migration	Caval Perforation	Caval Thrombosis	Pulmonary Embolism	Filter Fracture	Guidewire Entrapment	Misplaced Deployment	Partial Deployment	Failed Deployment	Tilted Filter	Other	Detachment of Component(s)
28	G2 Actual	0.000%	0.000%	0.1781%	0.7124%	0.8905%	0.0445%	0.0000%	0.0445%	0.1336%	0.000%	0.000%	0.0003%	0.000%	0.0233%	0.000%	0.0890%	0.000%	0.000%	0.000%
29	SNF	0.000%	0.000%	0.0335%	0.0322%	0.0657%	0.0013%	0.0013%	0.0028%	0.0090%	0.000%	0.000%	0.0026%	0.000%	0.0052%	0.0129%	0.0142%	0.000%	0.000%	0.0193%
30	Recovery	0.0547%	0.0374%	0.3395%	0.0633%	0.4026%	0.0777%	0.0547%	0.1324%	0.0374%	0.000%	0.0317%	0.1295%	0.000%	0.0058%	0.0345%	0.0230%	0.0029%	0.000%	0.0058%
31	Vena Tech	0.0078%	0.0059%	0.0624%	0.0410%	0.1034%	0.0078%	0.0449%	0.0927%	0.000%	0.000%	0.0020%	0.0023%	0.0020%	0.000%	0.0078%	0.0312%	0.000%	0.000%	0.0059%
32	Greenfield	0.0075%	0.0015%	0.0426%	0.1113%	0.1539%	0.0070%	0.0170%	0.0241%	0.0050%	0.000%	0.0020%	0.0045%	0.0090%	0.0208%	0.0652%	0.0196%	0.000%	0.000%	0.0060%
33	Bird's Nest	0.0138%	0.0039%	0.2897%	0.1242%	0.4138%	0.0138%	0.0138%	0.0276%	0.1517%	0.0039%	0.0138%	0.0993%	0.000%	0.0028%	0.000%	0.0414%	0.000%	0.000%	0.000%
34	Tulip	0.0138%	0.0039%	0.0770%	0.0691%	0.1481%	0.0138%	0.0158%	0.0276%	0.1517%	0.0039%	0.0138%	0.0993%	0.000%	0.0028%	0.000%	0.0414%	0.000%	0.000%	0.000%
35	TrapEase	0.0151%	0.0047%	0.0763%	0.0171%	0.0934%	0.0052%	0.0089%	0.0151%	0.0089%	0.0322%	0.0042%	0.0093%	0.000%	0.000%	0.0078%	0.0069%	0.000%	0.0057%	0.000%
36	OptEase	0.0177%	0.000%	0.1381%	0.0286%	0.1656%	0.0237%	0.000%	0.0237%	0.000%	0.0237%	0.0237%	0.0473%	0.000%	0.0118%	0.000%	0.0177%	0.000%	0.0177%	0.000%
37																				
38																				
39																				
40																				
41																				
42																				
43																				
44																				

DRAFT

Confidential & Privileged

No data available (based on last 4 quarters' trend)

[illegible]

[illegible]

A	B	C	D	E	F	G	H	I	J	K	L	M
Death	Migration	Filter Embolization	Caval Perforation	Caval Thrombosis	Pulmonary Embolism	Filter Fracture	Guidewire Entrapment	Misplaced Deployment	Partial Deployment	Failed Deployment	Tilted Filter	Other
1												
2	0	1	7	0	0	2	0	4	10	11	0	0
3												
4												
5												
6	Company	Device	Date	Source	FDA #	Failure	Death?	Filter Embolization	0			
7	BPV	SIMON NITINOL VENA	5/2/2005	KMJ	599074	Caval Perforation						
8	BPV	SIMON NITINOL FILTER	4/6/2005	KMJ	590761	Failed Deployment						
9	BPV	SIMON NITINOL FILTER	3/30/2005	KMJ	587919	Misplaced Deployment						
10	BPV	SIMON NITINOL FILTER	3/30/2005	KMJ	587882	N/A - Sales Rep issue						
11	BPV	SIMON NITINOL FILTER	3/28/2005	KMJ	586963	Partial Deployment						
12	BPV	SIMON NITINOL FILTER-F	3/16/2005	KMJ	584469	Partial Deployment	Filter & part of catheter implanted					
13	BPV	SIMON NITINOL FILTER	1/24/2005	KMJ	571075	Partial Deployment	"6th Foot" did not open					
14	BPV	SIMON NITINOL VENA C	1/5/2005	KMJ	565692	Partial Deployment	Twisted Legs					
15	BPV	SIMON NITINOL VENA C	11/29/04	BPV	557650	Failed Deployment	Crossed Legs					
16	BPV	SIMON NITINOL VENA C	11/15/04	BPV	558935	Failed Deployment						
17	BPV	SIMON NITINOL VENA C	08/20/04	BPV	541700	Misplaced Deployment	upside down deployment					
18	BPV	SIMON NITINOL VENA C	06/11/04	BPV	529838	Caval Perforation						
19	BPV	SIMON NITINOL FILTER	03/10/04	BPV	515514	Caval Perforation						
20	BPV	SIMON NITINOL FILTER	01/30/04	BPV	517796	Detachment of Component(s)	Marker Bands					
21	BPV	SIMON NITINOL FILTER	01/20/04	BPV	515079	Misplaced Deployment						
22	BPV	SIMON NITINOL FILTER	12/23/03	DJU	506286	Filter Fracture						
23	BPV	SIMON NITINOL FILTER	12/23/03	JWL	503017	Filter embolization	Leg in Renal Vein Migration (Rt Heart)					
24	BPV	SIMON NITINOL FILTER	12/12/03	BPV	504208	Detachment of Component(s)	Deployed Upside-down					
25	BPV	SIMON NITINOL VENA C	09/29/03	BPV	487339	Partial Deployment						
26	BPV	SIMON NITINOL FILTER	09/29/03	DJU	487339	Migration						
27	BPV	SIMON NITINOL FILTER	04/11/03	BPV	453965	Detachment of Component(s)	Movement 1 cm					
28	BPV	SIMON NITINOL FILTER	04/02/03	BPV	452201	Detachment of Component(s)	Marker Bands					
29	BPV	SIMON NITINOL FILTER	03/13/03	BPV	448182	Detachment of Component(s)	Marker Bands					
30	BPV	2320A-S/SNF ANTECUBI	01/17/03	BPV	438877	Detachment of Component(s)	Marker Bands					
31	BPV	2120F-S/SNF FEMORAL	11/18/02	BPV	429497	Failed Deployment	Legs Twisted					
32	BPV	2120F-S/SNF FEMORAL	11/18/02	BPV	429330	Failed Deployment	Legs Twisted					
33	BPV	2120F-S/SNF FEMORAL	11/18/02	BPV	429328	Failed Deployment	Legs Twisted					
34	BPV	2220J-S/SNF FEMORAL	11/08/02	BPV	428025	Filter Fracture	Filter Fracture					
35	BPV	2120F-S/SNF FEMORAL	08/20/02	BPV	412957	Partial Deployment	Legs did not open					
36	BPV	3220J-NITINOL JUGULAR	07/18/02	BPV	406553	Detachment of Component(s)	Marker Bands					
37	BPV	SIMON NITINOL FILTER	06/21/02	BPV	401126	Detachment of Component(s)	Marker Bands					
38	BPV	SIMON NITINOL FILTER	06/17/02	BPV	400874	Caval Perforation	Marker Bands					
39	BPV	SIMON NITINOL FILTER	05/22/02	DJU	396361	Detachment of Component(s)	Marker Bands					
40	BPV	SIMON NITINOL FILTER	05/13/02	BPV	394027	Detachment of Component(s)	Marker Bands					
41	BPV	SIMON NITINOL FILTER	04/15/02	BPV	388439	Caval Perforation	Marker Bands					
42	BPV	SIMON NITINOL FILTER	04/02/02	BPV	386983	Detachment of Component(s)	Marker Bands					
43	BPV	SIMON NITINOL FILTER	03/22/02	BPV	384797	Caval Perforation	Marker Bands					
44	BPV	SIMON NITINOL FILTER	03/05/02	BPV	381558	Failed Deployment	Hooks Stuck					
45	BPV	SIMON NITINOL FILTER	03/05/02	BPV	380816	Caval Perforation	Hooks Stuck					
46	BPV	SIMON NITINOL FILTER	02/11/02	BPV	377174	Failed Deployment	Hooks Stuck					
47	BPV	SIMON NITINOL FILTER	02/11/02	DJU	377169	Misplaced Deployment	Hooks Stuck					

[illegible]

	A	B	C	D	E	F	G	H	I	J
	Death	Migration	Filter Embolization	Caval Perforation	Caval Thrombosis	Pulmonary Embolism	Filter Fracture	Guidewire Entrapment	Misplaced Deployment	Partial Deployment
1	19	27	19	13	0	11	45	0	2	12
2										
3										
4										
5										
6	Company	Device	Date	Source	FDA #	Failure	Comments	Death	Filter Emb death?	13
7	BPV	BARD	9/18/2005	KMJ	636272	Caval Perforation	may be duplicate of 636277			
8	BPV	RECOVERY FILTER	9/13/2005	KMJ	636277	Caval Perforation				
9	BPV	RECOVERY FILTER	9/12/2005	KMJ	634437	Pulmonary Embolism	moved up "one vertebra"			
10	BPV	RECOVERY FILTER	9/8/2005	KMJ	633930	Migration	4 cm to iliac vein			
11	BPV	RECOVERY FILTER	9/7/2005	KMJ	633727	Migration	to heart @ implant			
12	BPV	RECOVERY FILTER	9/1/2005	KMJ	633205	Filter Fracture	633205: Migr & 1 Detached Leg			
13	BPV	RECOVERY FILTER	9/1/2005	KMJ	633205	Filter Embolization	to tricuspid valve w/ heavy clot burden	Death	1	
14	BPV	RECOVERY VENA CAVA F	8/24/2005	KMJ	631194	Migration	to tricuspid valve w/ clot			
15	BPV	RECOVERY FILTER	8/9/2005	KMJ	627651	Filter Fracture	Detached partial leg or hook?			
16	BPV	RECOVERY FILTER	8/1/2005	KMJ	624371	Failed Deployment				
17	BPV	RECOVERY FILTER	7/29/2005	KMJ	631071	Misplaced Deployment				
18	BPV	RECOVERY FILTER	7/27/2005	KMJ	629054	Filter Fracture	1 Arm - caught in filter & removed			
19	BPV	RECOVERY FILTER	7/14/2005	KMJ	621545	Migration	to heart w/ clot burden			
20	BPV	RECOVERY FILTER	7/14/2005	KMJ	621543	Filter Fracture	2 Arms			
21	BPV	BARD SYSTEMS	7/12/2005	KMJ	621074	Filter Fracture = 619697	duplicate			
22	BPV	RECOVERY FILTER	7/12/2005	KMJ	621078	Filter Fracture	1 Arm in lung			
23	BPV	BARD RECOVERY FILT	7/8/2005	KMJ	620484	Migration	to heart w/ multiple clots			
24	BPV	RECOVERY FILTER	7/7/2005	KMJ	620888	Filter Fracture	1 Arm in cava wall - removed successfully			
25	BPV	RECOVERY FILTER	7/7/2005	KMJ	620897	Filter Fracture	1 Arm & 2 Hooks			
26	BPV	RECOVERY FILTER	7/7/2005	KMJ	620859	Filter Fracture	2 Arms (rt atrium & pulm artery)			
27	BPV	BARD SYSTEMS	7/1/2005	KMJ	619697	Filter Fracture	1 Arm in lower lobe/pulm artery branch			
28	BPV	RECOVERY FILTER	7/1/2005	KMJ	619081	Migration	Hepatic/right atrium junction			
29	BPV	RECOVERY FILTER	7/1/2005	KMJ	616355	Migration	Rt atrium w/ clot burden			
30	BPV	RECOVERY FILTER	6/16/2005	KMJ	615340	Filter Embolization	Filter & clot in right atrium	Death	1	
31	BPV	RECOVERY FILTER	6/15/2005	KMJ	614686	Filter Embolization	Both pulm arteries, filter w/massive clot	Death	1	
32	BPV	RECOVERY FILTER	6/15/2005	KMJ	614668	Migration	5cm into the hepatic ivc			
33	BPV	RECOVERY FILTER	6/13/2005	KMJ	614127	Filter Embolization	Filter & clot in rt atrium	Death	1	
34	BPV	BARD RECOVERY	6/13/2005	KMJ	614168	Migration				
35	BPV	RECOVERY FILTER	6/3/2005	KMJ	611395	Pulmonary Embolism				
36	BPV	RECOVERY FILTER	6/3/2005	KMJ	618010	Filter Fracture				
37	BPV	BARD RECOVERY FILTER	5/26/2005	KMJ	608450	Filter Fracture	3cm into Renals + 4 arms detached			
38	BPV	BARD RECOVERY FILTER	5/26/2005	KMJ	608450	Migration	3cm into Renals + 4 arms detached			
39	BPV	RECOVERY FILTER	5/23/2005	KMJ	597752	Duplicate	Migration - Distal the right atrium			
40	BPV	RECOVERY FILTER	5/10/2005	KMJ	603617	Filter Fracture	1 hook missing			
41	BPV	BARD RECOVERY FILTER	5/3/2005	KMJ	598730	Filter Fracture	1 arm in right pulmonary artery			
42	BPV	RECOVERY FILTER	4/26/2005	KMJ	597752	Migration	Distal the right atrium			
43	BPV	RECOVERY FILTER	4/26/2005	KMJ	598730	Filter Fracture	1 arm in right pulmonary artery			
44	BPV	BARD IVC FILTER	4/14/2005	KMJ	593418	Duplicate	Duplicate in database			
45	BPV	BARD IVC FILTER	4/14/2005	KMJ	593418	Filter Fracture				
46	BPV	RECOVERY FILTER	4/13/2005	KMJ	592511	Partial Deployment	Twisted legs			
47	BPV	RECOVERY FILTER	4/13/2005	KMJ	592547	Filter Fracture	Limb in lung			
48	BPV	RECOVERY FILTER	4/12/2005	KMJ	592748	Filter Fracture	2 arms - renal vein & iliac			
49	BPV	RECOVERY FILTER	4/12/2005	KMJ	591237	Migration	Tip in right atrium			
50	BPV	RECOVERY FILTER	4/12/2005	KMJ	592730	Failed Deployment				

A	B	C	D	E	F	G	H	I	J
51 BPV	RECOVERY FILTER	4/12/2005	KMJ	592740	Caval Perforation				
52 BPV	RECOVERY FILTER	4/7/2005	KMJ	590300	Caval Perforation				
53 BPV	RECOVERY FILTER	4/4/2005	KMJ	588407	Pulmonary Embolism		Death		
54 BPV	RECOVERY FILTER	3/25/2005	KMJ	587269	Filter Fracture				
55 BPV	RECOVERY FILTER	3/25/2005	KMJ	569852	Tilted Filter				
56 BPV	RECOVERY FILTER	3/15/2005	KMJ	585163	Partial Deployment				
57 BPV	RECOVERY FILTER	3/15/2005	KMJ	575175	Filter Fracture	Twisted Legs			
58 BPV	RECOVERY FILTER	3/14/2005	KMJ	584484	Filter Fracture				
59 BPV	BARD	3/14/2005	KMJ	584473	Filter Fracture				
60 BPV	BARD RECOVERY (VC F)	3/14/2005	KMJ	581678	Filter Fracture				
61 BPV	RECOVERY FILTER	3/10/2005	KMJ	583349	Partial Deployment				
62 BPV	RECOVERY FILTER	3/7/2005	KMJ	580296	Pulmonary Embolism		Death		
63 BPV	RECOVERY FILTER	3/4/2005	KMJ	582383	Filter Fracture				
64 BPV	RECOVERY FILTER	3/4/2005	KMJ	582362	Filter Fracture				
65 BPV	RECOVERY FILTER	3/3/2005	KMJ	581426	Misplaced Deployment	accessory vein			
66 BPV	RECOVERY FILTER	2/25/2005	KMJ	580062	Filter Fracture				
67 BPV	RECOVERY FILTER	2/25/2005	KMJ	578634	Pulmonary Embolism				
68 BPV	BARD	2/15/2005	KMJ	575244	= 567210	Duplicate of 567210	Death		
69 BPV	RECOVERY FILTER	2/9/2005	KMJ	572544	Pulmonary Embolism				
70 BPV	RECOVERY FILTER	2/8/2005	KMJ	574995	Migration	5 cm movement w/ clot	Death		
71 BPV	RECOVERY FILTER	2/8/2005	KMJ	572279	Migration	right atrium			
72 BPV	UNKNOWN BRAND NAME	1/31/2005	KMJ	584859	Filter Fracture				
73 BPV	RECOVERY FILTER	1/31/2005	KMJ	574154	Filter Fracture	Fracture + Migration			
74 BPV	RECOVERY FILTER	1/31/2005	KMJ	574154	Filter Fracture	Fracture + Migration			
75 BPV	BARD RECOVERY FILTER	1/31/2005	KMJ	570337	Filter Fracture				
76 BPV	RECOVERY FILTER	1/31/2005	KMJ	569583	Migration	3 cm movement w/ clot			
77 BPV	RECOVERY FILTER	1/28/2005	KMJ	571941	Migration	2 cm movement w/ 10cm of clot			
78 BPV	RECOVERY FILTER	1/24/2005	KMJ	571084	Filter Fracture	legs opened asymmetrically			
79 BPV	RECOVERY FILTER	1/19/2005	KMJ	569799	Partial Deployment				
80 BPV	RECOVERY FILTER	1/18/2005	KMJ	570361	Filter Fracture				
81 BPV	RECOVERY FILTER	1/14/2005	KMJ	569850	Filter Fracture				
82 BPV	RECOVERY FILTER	1/14/2005	KMJ	569848	Partial Deployment	Twisted Legs			
83 BPV	RECOVERY FILTER	1/13/2005	KMJ	587274	Filter Fracture				
84 BPV	BARD	1/13/2005	KMJ	567210	Filter Fracture	arm in pulmonary artery at autopsy	Death		
85 BPV	RECOVERY FILTER	1/12/2005	KMJ	568356	Filter Fracture				
86 BPV	RECOVERY FILTER	1/6/2005	KMJ	565461	Failed Deployment	Arms and Legs did not open			
87 BPV	VENA CAVA FILTER	1/6/2005	KMJ	564519	= 567210	Duplicate of 567210			
88 BPV	RECOVERY FILTER	1/5/2005	KMJ	565157	Failed Deployment	Legs did not open			
89 BPV	RECOVERY FILTER	1/5/2005	KMJ	564389	Filter Fracture				
90 BPV	RECOVERY FILTER	1/5/2005	KMJ	564354	Pulmonary Embolism				
91 BPV	RECOVERY FILTER	12/22/2004	BPV	565413		Duplicate			
92 BPV	RECOVERY FILTER	12/22/2004	BPV	561581	Caval Perforation				
93 BPV	RECOVERY FILTER	12/21/2004	BPV	561390	Migration				
94 BPV	RECOVERY FILTER	12/15/2004	BPV	560531	Filter Embolization				
95 BPV	RECOVERY FILTER	12/14/2004	BPV	560255	Partial Deployment	Twisted Legs	Death	1	
96 BPV	RECOVERY FILTER	12/14/2004	BPV	560255	Partial Deployment	Twisted Legs			
97 BPV	BARD RECOVERY FILTER	12/9/2004	BPV	562812	Duplicate				
98 BPV	BARD RECOVERY FILTER	12/9/2004	BPV	562817	Filter Embolization				
99 BPV	RECOVERY FILTER	12/8/2004	BPV	558963	Partial Deployment	Twisted Legs			
100 BPV	RETRIEVABLE GREENFIE	12/7/2004	BPV	560997	Filter Fracture				
101 BPV	RECOVERY FILTER	12/1/2004	BPV	557639	Caval Perforation				
102 BPV	RECOVERY FILTER	11/10/2004	BPV	554903	Migration	T-12			

A	B	C	D	E	F	G	H	I	J
103 BPV	BARD RECOVERY FILTER	11/4/2004	BPV	557688	Filter Embolization		Death		
104 BPV	RECOVERY FILTER	10/26/2004	BPV	552526	Filter Fracture				
105 BPV	RECOVERY FILTER	10/14/2004	BPV	549338	Filter Embolization	Rt Heart	Death	1	
106 BPV	RECOVERY FILTER	10/14/2004	BPV	549017	Filter Embolization	Heart		1	
107 BPV	RECOVERY FILTER	10/14/2004	BPV	549017	Failed Deployment	Did not Open			
108 BPV	BARD RECOVERY IVC FI	10/5/2004	BPV	557154	Migration	Caudal - 4cm			
109 BPV	BARD RECOVERY IVC FI	10/5/2004	BPV	557154	Filter Fracture	Arm			
110 BPV	BARD RECOVERY IVC FI	10/5/2004	BPV	556875		Same as 557154			
111 BPV	RECOVERY FILTER	10/1/2004	BPV	547044	Migration				
112 BPV	RECOVERY FILTER	10/1/2004	BPV	547044	Filter Fracture	Hook			
113 BPV	RECOVERY FILTER	9/21/2004	JWL	545303	Filter Embolization	intrahepatic IVC	Death	1	
114 BPV	RECOVERY FILTER	9/21/2004	BPV	545298	Caval Perforation				
115 BPV	RECOVERY FILTER	9/21/2004	JWL	545283	Caval Perforation	Removed, fine			
116 BPV	RECOVERY FILTER	9/10/2004	BPV	543326	Caval Perforation				
117 BPV	RECOVERY FILTER	9/9/2004	BPV	543154	Failed Deployment	Crossed Legs			
118 BPV	RECOVERY FILTER	9/9/2004	BPV	543080	= 545130				
119 BPV	RECOVERY FILTER	9/9/2004	BPV	543075	Failed Deployment	Crossed Legs			
120 BPV	RECOVERY FILTER	9/9/2004	BPV	543065	Pulmonary embolism	6-8cm			
121 BPV	RECOVERY FILTER	9/9/2004	BPV	543065	Migration	6-8cm			
122 BPV	RECOVERY FILTER	8/30/2004	BPV	541694	Other	Pain			
123 BPV	RECOVERY FILTER	8/30/2004	JWL	541690	Filter embolization	right atrium			
124 BPV	RECOVERY FILTER	8/26/2004	JWL	541353	Filter embolization	heart	Death	1	
125 BPV	RECOVERY FILTER	8/23/2004	JWL	540532	Filter embolization	heart, placed into large cava			
126 BPV	RECOVERY FILTER	8/19/2004	BPV	540258	Caval Perforation				
127 BPV	BARD-RECOVERY FILTER	8/13/2004	BPV	545130	Filter Fracture				
128 BPV	RECOVERY FILTER	8/5/2004	BPV	538204	migration	L3 to T9/10			
129 BPV	RECOVERY FILTER	8/5/2004	BPV	538201	migration	L2 to T12			
130 BPV	RECOVERY FILTER	8/4/2004	BPV	538046	Caval perforation				
131 BPV	FEMORAL RECOVERY FIL	7/30/2004	JWL	538589	= 541353	Heart			
132 BPV	RECOVERY FILTER	7/28/2004	BPV	537411	Filter Fracture				
133 BPV	RECOVERY FILTER	7/14/2004	JWL	534603	= 534597	duplicate of 534597			
134 BPV	BARD	7/14/2004	JWL	534597	Filter embolization	right ventricle	Death	1	
135 BPV	RECOVERY FILTER	7/7/2004	BPV	533379	Other	Pain			
136 BPV	RECOVERY FILTER	6/21/2004	BPV	541295	Filter Fracture	retrieval attempts caused fracture?			
137 BPV	RECOVERY FILTER	6/21/2004	JWL	531130	Filter embolization	tricuspid valve			
138 BPV	RECOVERY FILTER	6/14/2004	JWL	530079	Filter embolization	heart	Death	1	
139 BPV	RECOVERY FILTER	6/4/2004	JWL	531446	= 525383	duplicate of 525383			
140 BPV	BARD PERIPHERAL VASC	6/1/2004	JWL	528008	= 530079	duplicate of 530079			
141 BPV	RECOVERY VENA CAVA	5/24/2004	BPV	526245	Filter Fracture	4cm + PE			
142 BPV	RECOVERY FILTER	5/21/2004	BPV	526110	Pulmonary embolism	4cm + PE			
143 BPV	RECOVERY FILTER	5/21/2004	BPV	526110	Migration	right ventricle	Death	1	
144 BPV	RECOVERY FILTER	5/12/2004	JWL	525383	Filter embolization	duplicate of 525383			
145 BPV	CR BARD	4/28/2004	JWL	522655	= 525383				
146 BPV	RECOVERY FILTER	4/16/2004	BPV	521172	Caval perforation				
147 BPV	RECOVERY FILTER	3/31/2004	BPV	542132	Partial Deployment	Crossed Legs			
148 BPV	RECOVERY FILTER	3/31/2004	BPV	541621	Partial Deployment				
149 BPV	RECOVERY FILTER	3/18/2004	BPV	536423	Filter Fracture				
150 BPV	RECOVERY FILTER	3/18/2004	BPV	516863	Filter embolization	right atrium/doctor pushed it there			
151 BPV	BARD RECOVERY FILTER	3/8/2004	JWL	515237	Filter embolization	right atrium	Death	1	
152 BPV	RECOVERY FILTER	3/2/2004	BPV	514355	Pulmonary Embolism	Filter tilted			
153 BPV	BARD	2/17/2004	JWL	511801	= 515237	duplicate of 515237			
154 BPV	RECOVERY FILTER	2/15/2004	BPV	513064	Partial Deployment	Crossed Legs			

[illegible]

	A	B	C	D	E	F	G	H	I
	Death	Migration	Filter Embolization	Caval Perforation	Caval Thrombosis	Pulmonary Embolism	Filter Fracture	Guidewire Entrapment	Misplaced Deployment
1	4	4	23	0	0	1	1	1	0
2									
3									
4									
5									
6	Company	Device	Date	Source	FDA #	Failure	Location	Death?	Filter Emb death?
7	B. BRAUN MEDICAL	VENA TECH	8/19/2005	KMJ	628618	Migration	Did not open at implant, migrated to heart/lungs		
8	B. BRAUN MEDICAL	VENA TECH	8/5/2005	KMJ	624755	Failed Deployment	Legs would not open		
9	B. BRAUN MEDICAL	VENA TECH	7/27/2005	KMJ	623284	Failed Deployment	Did not open		
10	B. BRAUN MEDICAL, IN	VENA TECH	6/24/2005	KMJ	617223	Failed Deployment	Deployed in pt's gonadel vein		
11	B. BRAUN MEDICAL, IN	VENA TECH	6/23/2005	KMJ	617110	Failed Deployment	Filter then turned upside down		
12	B. BRAUN MEDICAL, IN	VENA TECH	6/13/2005	KMJ	614111	Failed Deployment			
13	B. BRAUN MEDICAL, IN	VENA TECH	6/13/2005	KMJ	613670	Filter embolization	Pulmonary Artery	Death	1
14	B. BRAUN MEDICAL, IN	VENA TECH	6/8/2005	KMJ	612832	Failed Deployment	Did not open & then tilted		
15	B. BRAUN MEDICAL, IN	VENA TECH	5/31/2005	KMJ	610574	Duplicate = 586485			
16	B. BRAUN MEDICAL, IN	VENA-TECH LP	4/22/2005	KMJ	586302	Failed Deployment	Surgery required to remove filter		
17	B. BRAUN MEDICAL, IN	BRAUN VENA-TECH FILT	3/23/2005	KMJ	586485	Migration	Filter & mass in rt atrium		
18	B. BRAUN MEDICAL, IN	VENA TECH	12/03/04	BPV	557804	Filter embolization	RA		
19	B. BRAUN MEDICAL, IN	VENA TECH	10/07/04	BPV	548359	Filter Tilt			
20	B. BRAUN MEDICAL, IN	VENA TECH	10/07/04	BPV	548359	Migration			
21	B. BRAUN MEDICAL, INC	VENA TECH	08/19/04	BPV	540178	guidewire entrapment			
22	B. BRAUN MEDICAL, INC	VENA TECH	08/02/04	JWL	539352	Filter embolization	PA		
23	B. BRAUN MEDICAL, INC	B. BRAUN	06/29/04	JWL	537548	Filter embolization	PA		
24	B. BRAUN MEDICAL, INC	VENA TECH	06/16/04	JWL	530752	Partial deployment			
25	B. BRAUN MEDICAL, INC	VENA TECH LP	05/14/04	JWL	525261	Failed Deployment			
26	B. BRAUN MEDICAL, INC	VENA TECH	01/08/04	BPV	505702	Failed Deployment			
27	B. BRAUN MEDICAL, INC	VENA TECH LP	12/24/03	JWL	502789	Filter embolization	Migration (RA)/Partial Deployment		
28	B. BRAUN MEDICAL, INC	VENA CAVA FILTER	12/23/03	BPV	503348	Other	Migration (PA)		
29	B. BRAUN MEDICAL, INC	VENA TECH IVC FILTER	12/03/03	JWL	502269	Filter embolization			
30	B. BRAUN MEDICAL, INC	VENA TECH LP	10/09/03	BPV	488892	Failed Deployment	Migration (Lung)/Partial Deployment		
31	B. BRAUN MEDICAL, INC	VENA CAVA FILTER SYS	10/03/03	JWL	487738	Filter embolization	Migration (Pulmonary Artery)/Partial Deployment		
32	B. BRAUN MEDICAL, INC	VENA TECH	05/12/03	JWL	459882	Filter embolization	Migration (Pulmonary Valve)/Partial Deployment	Death	1
33	B. BRAUN MEDICAL, INC	VENA TECH LP	04/08/03	JWL	454732	= 459882	Migration (Suprarenal)		
34	B. BRAUN MEDICAL, INC	VENA TECH LP	02/18/03	BPV	443744	Migration			
35	B. BRAUN MEDICAL, INC	VENA TECH	02/10/03	BPV	443031	Filter Fracture			
36	B. BRAUN MEDICAL, INC	VENA TECH LP	01/24/03	BPV	439121	Partial Deployment			
37	B. BRAUN MEDICAL, INC	VENA TECH	12/16/02	JWL	433349	Filter embolization	Migration (PA)		
38	B. BRAUN MEDICAL, INC	VENA TECH	12/09/02	JWL	431971	Filter embolization	Migration (Rt Atrium)		
39	B. BRAUN MEDICAL, INC	VENA TECH	11/22/02	BPV	429650	Failed Deployment			
40	B. BRAUN MEDICAL, INC	VENA TECH	11/22/02	JWL	429648	Filter embolization	Migration (PA)		
41	B. BRAUN MEDICAL, INC	VENA TECH LP	05/20/02	JWL	395792	Filter embolization	Migration (Rt Atrium)		
42	B. BRAUN MEDICAL, INC	VENA TECH	05/17/02	JWL	397413	Filter embolization	Migration (PA)		
43	B. BRAUN MEDICAL, INC	VENA TECH LP	05/17/02	JWL	395251	Filter embolization	Migration (Intrahepatic)		
44	B. BRAUN MEDICAL, INC	VENA TECH	05/14/02	JWL	394714	Filter embolization	Migration (PA)		
45	B. BRAUN MEDICAL, INC	VENA TECH	05/07/02	JWL	393904	Filter embolization	Migration (PA)		
46	B. BRAUN MEDICAL, INC	VENA TECH	04/09/02	JWL	388359	Filter embolization	Migration (Pulmonary Valve)		
47	B. BRAUN MEDICAL, INC	1 VENA TECH LP	04/03/02	JWL	385985	Filter embolization	Migration (heart)		
48	B. BRAUN MEDICAL, INC	VENA TECH	02/11/02	BPV	377437	Partial Deployment			

[illegible]

	A	B	C	D	E	F	G	H	I
	Death	Migration	Filter Embolization	Caval Perforation	Caval Thrombosis	Pulmonary Embolism	Filter Fracture	Guidewire Entrapment	Misplaced Deployment
1									
2	15	14	34	10	1	4	9	12	41
3									
4									
5									
6	Company	Device	Date	Source	FDA #	Failure	Notes	Death?	Filter Emb death?
7	BOST SCI	TITANIUM GREENFIELD	9/7/2005	KMJ	633797	Caval Perforation	5 legs perforated caval wall at implant		
8	BOST SCI	GREENFIELD	9/2/2005	KMJ	636383	Migration	Legs crossed		
9	BOST SCI	TITANIUM GREENFIELD	8/31/2005	KMJ	632439	Partial Deployment			
10	BOST SCI	GREENFIELD 12 FR SS	7/28/2005	KMJ	623432	Guidewire Entrapment			
11	BOST SCI	GREENFIELD 12 FR SS	5/26/2005	KMJ	608237	Migration	to heart		
12	BOST SCI	TITANIUM GREENFIELD	5/26/2005	KMJ	608191	Failed Deployment	Deployed prematurely, pt required surgery		
13	BOST SCI	TITANIUM GREENFIELD	5/26/2005	KMJ	611695	Failed Deployment			
14	BOST SCI	GREENFIELD	5/23/2005	KMJ	609363	Misplaced Deployment	Right iliac vein		
15	BOST SCI	UNKNOWN BRAND NAME	5/16/2005	KMJ	606333	Migration	Rt ventricle then PA & Bard Recovery filter placed		
16	BOST SCI	TITANIUM GREENFIELD	5/13/2005	KMJ	603559	Failed Deployment			
17	BOST SCI	TITANIUM GREENFIELD	5/13/2005	KMJ	604240	Partial Deployment			
18	BOST SCI	GREENFIELD 12 FR SS	5/6/2005	KMJ	601746	Failed Deployment	Deployed outside of pt		
19	BOST SCI	TITANIUM GREENFIELD	4/14/2005	KMJ	592448	Migration	Left pulmonary artery		
20	BOST SCI	GREENFIELD 12 FR SS	4/6/2005	KMJ	568929	Failed Deployment			
21	BOST SCI	GREENFIELD 12 FR SS	4/4/2005	KMJ	589053	Other	Packaging damaged - pin hole		
22	BOST SCI	GREENFIELD 12 FR SS	3/18/2005	KMJ	584666	Misplaced Deployment	Filter deployed early		
23	BOST SCI	TITANIUM GREENFIELD	3/17/2005	KMJ	585352	Failed Deployment	2 filters failed in same pt		
24	BOST SCI	TITANIUM GREENFIELD	3/17/2005	KMJ	585136	Failed Deployment	3rd worked, pt expired	Death	
25	BOST SCI	TITANIUM GREENFIELD	3/17/2005	KMJ	585153	Partial Deployment			
26	BOST SCI	TITANIUM GREENFIELD	3/16/2005	KMJ	582601	Partial Deployment	Partial Deployment + Migration		
27	BOST SCI	TITANIUM GREENFIELD	3/16/2005	KMJ	582601	Migration	Partial Deployment + Migration		
28	BOST SCI	GREENFIELD 12 FR SS	3/10/2005	KMJ	580668	Misplaced Deployment	Filter deployed early		
29	BOST SCI	GREENFIELD 12 FR SS	3/2/2005	KMJ	577957	Other	Bleeding		
30	BOST SCI	GREENFIELD 12 FR SS	3/2/2005	KMJ	577953	Other	Duplicate of 577957		
31	BOST SCI	GREENFIELD VENA CAVA	2/25/2005	KMJ	577322	Failed Deployment			
32	BOST SCI	TITANIUM GREENFIELD	2/18/2005	KMJ	574621	Partial Deployment			
33	BOST SCI	TITANIUM GREENFIELD	2/18/2005	KMJ	574656	Partial Deployment	Duplicate of 574621		
34	BOST SCI	TITANIUM GREENFIELD	2/18/2005	KMJ	574644	Partial Deployment	Duplicate of 574621		
35	BOST SCI	TITANIUM GREENFIELD	2/18/2005	KMJ	574665	Partial Deployment	Duplicate of 574621		
36	BOST SCI	GREENFIELD 12 FR SS	2/11/2005	KMJ	572385	Failed Deployment	Deployed during pre-op		
37	BOST SCI	FILTERWIRE EX EMBOLI	1/31/2005	KMJ	574156	Failed Deployment			
38	BOST SCI	TITANIUM GREENFIELD	1/26/2005	KMJ	567967	Failed Deployment			
39	BOST SCI	GREENFIELD 12 FR VEN	1/20/2005	KMJ	566637	Partial Deployment	Twisted Legs		
40	BOST SCI	GREENFIELD 12 FR SS	1/7/2005	KMJ	564587	Partial Deployment			
41	BOST SCI	TITANIUM GREENFIELD	1/7/2005	KMJ	564584	Failed Deployment	2 filters failed in same pt		
42	BOST SCI	TITANIUM GREENFIELD	1/7/2005	KMJ	564581	Failed Deployment	(not duplicate)		
43	BOST SCI	GREENFIELD 12 FR SS	1/5/2005	KMJ	563939	Partial Deployment			
44	BOST SCI	GREENFIELD 12FR SS V	12/23/04	BPV		Failed Deployment			
45	BOST SCI	GREENFIELD 12 FR SS	12/14/04	BPV		Filter embolization	RV	Death	1
46	BOST SCI	GREENFIELD VENACAVA	12/13/04	BPV		Filter embolization	RA		
47	BOST SCI	TITANIUM GREENFIELD	12/10/04	BPV		Misplaced Deployment			

A	B	C	D	E	F	G	H	I
48	BOST SCI GREENFIELD 12 FR SS	12/08/04	BPV		Filter Fracture			
49	BOST SCI TITANIUM GREENFIELD	12/02/04	BPV		Partial deployment			
50	BOST SCI GREENFIELD 12 FR SS	11/24/04	BPV		Failed deployment			
51	BOST SCI GREENFIELD 12 FR SS	11/23/04	BPV	556832	Filter embolization	Heart	Death	1
52	BOST SCI 12 FR TITANIUM GREEN	11/19/04	BPV		Partial deployment			
53	BOST SCI GREENFIELD 12 FR SS	11/19/04	BPV		Failed deployment			
54	BOST SCI GREENFIELD 12 FR SS	11/10/04	BPV		Failed deployment			
55	BOST SCI GREENFIELD 12 FR SS	11/05/04	BPV		Misplaced Deployment			
56	BOST SCI GREENFIELD 12 FR SS	11/05/04	BPV	553901	Failed deployment			
57	BOST SCI GREENFIELD VENA CAVA	11/02/04	BPV		Failed deployment			
58	BOST SCI TITANIUM GREENFIELD	10/22/04	BPV		Failed deployment			
59	BOST SCI TITANIUM GREENFIELD	10/13/04	BPV		Partial deployment			
60	BOST SCI TITANIUM GREENFIELD	10/05/04	BPV		Partial deployment			
61	BOST SCI GREENFIELD 12 FR SS	10/04/04	BPV		Filter embolization	RA		
62	BOST SCI TITANIUM GREENFIELD	09/27/04	BPV	546066	Partial Deployment	did not open		
63	BOST SCI GREENFIELD 12 FR SS	09/17/04	JWL	544824	Filter embolization	did not open	Death	1
64	BOST SCI BOSTON SCIENTIFIC	09/14/04	BPV	544555	Misplaced Deployment			
65	BOST SCI TITANIUM GREENFIELD	09/02/04	BPV	541936	Partial Deployment	did not open		
66	BOST SCI GREENFIELD 12 FR SS	08/18/04	BPV	542415	Partial deployment	Crossed Legs		
67	BOST SCI TITANIUM GREENFIELD	08/13/04	BPV	539718	Partial Deployment			
68	BOST SCI GREENFIELD 12 FR SS	08/13/04	BPV	539702	Partial deployment	did not open		
69	BOST SCI GREENFIELD 12 FR SS	07/29/04	BPV	539670	Partial deployment			
70	BOST SCI GREENFIELD 12 FR SS	07/23/04	BPV	539747	Other	out of box failure		
71	BOST SCI TITANIUM GREENFIELD	07/22/04	BPV	539743	Partial Deployment			
72	BOST SCI GREENFIELD 12 FR SS	07/02/04	BPV	533037	Failed Deployment			
73	BOST SCI GREENFIELD 12 FR SS	06/25/04	BPV	539659	Mislabel	mislabeling		
74	BOST SCI GREENFIELD 12 FR SS	06/25/04	JWL	539661	= 539659			
75	BOST SCI GREENFIELD 12 FR SS	06/25/04	JWL	539660	= 539659			
76	BOST SCI TITANIUM GREENFIELD	06/11/04	BPV	539700	Other	out of box failure		
77	BOST SCI GREENFIELD 12FR SS V	06/07/04	BPV	539486	Other	out of box failure		
78	BOST SCI GREENFIELD 12FR SS V	05/28/04	BPV	539436	Other	out of box failure		
79	BOST SCI GREENFIELD 12 FR SS	05/28/04	JWL	539484	Mislabel			
80	BOST SCI GREENFIELD 12 FR SS	05/28/04	BPV	539474	Mislabel	mislabeling		
81	BOST SCI GREENFIELD VENA CAVA	05/28/04	BPV	542507	Filter Fracture			
82	BOST SCI TITANIUM GREENFIELD	05/26/04	BPV	527119	Failed Deployment	119 and 115 are for 2 GFs		
83	BOST SCI TITANIUM GREENFIELD	05/26/04	JWL	527115	Failed Deployment	119 and 115 are for 2 GFs		
84	BOST SCI GREENFIELD FILTER	05/26/04	BPV	528700	Caval Perforation			
85	BOST SCI TITANIUM GREENFIELD	05/21/04	BPV	539495	Failed Deployment			
86	BOST SCI TITANIUM GREENFIELD	05/21/04	JWL	539503	= 539495			
87	BOST SCI TITANIUM GREENFIELD	05/21/04	JWL	539489	= 539495			
88	BOST SCI GREENFIELD FILTER	05/14/04	BPV	543782	Caval Thrombosis			
89	BOST SCI GREENFIELD 12 FR SS	05/13/04	BPV	539510	Mislabel	mislabeling		
90	BOST SCI TITANIUM GREENFIELD	05/05/04	BPV	539523	Failed Deployment	= 539523, but 2 GFs involved		
91	BOST SCI TITANIUM GREENFIELD	05/05/04	JWL	539517	Failed Deployment	did not open		
92	BOST SCI GREENFIELD 12FR SS	05/04/04	JWL	526935	Migration			
93	BOST SCI TITANIUM GREENFIELD	04/23/04	BPV	525696	Partial Deployment			
94	BOST SCI TITANIUM GREENFIELD	04/16/04	BPV	539490	Partial Deployment	did not open		
95	BOST SCI GREENFIELD VENA CAVA	04/07/04	BPV	519607	Other			
96	BOST SCI GREENFIELD 12 FR SS	03/18/04	BPV	515350	Other	out of box failure		
97	BOST SCI TITANIUM GREENFIELD	03/11/04	BPV	515392	Migration			
98	BOST SCI GREENFIELD 12FR SS	03/05/04	JWL	514387	Migration	Migration, Did not open		
99	BOST SCI GREENFIELD 12FR SS V	02/20/04	BPV	525849	Partial Deployment	did not open		

A	B	C	D	E	F	G	H	I
100	BOST SCI GREENFIELD 12FR SS V	02/19/04	BPV	524148	Partial Deployment	did not open		
101	BOST SCI GREENFIELD 12FR SS	02/18/04	BPV	512053	Failed Deployment			
102	BOST SCI GREENFIELD VENA CAVA	02/06/04	BPV	520736	Partial Deployment			
103	BOST SCI TITANIUM GREENFIELD	02/05/04	BPV	518401	Misplaced deployment			
104	BOST SCI GREENFIELD FILTER VC	02/02/04	JWL	510364	Failed Deployment	Dual MDRs with same #		
105	BOST SCI TITANIUM GREENFIELD	01/28/04	BPV	516462	Failed Deployment	Stuck in Sheath		
106	BOST SCI GREENFIELD VENA CAVA	01/26/04	BPV	515533	Partial deployment	Crossed Legs		
107	BOST SCI GREENFIELD FILTER	12/11/03	JWL	502077	Guidewire problem	Guidewire Fracture		
108	BOST SCI GREENFIELD VENA CAVA	12/11/03	BPV	503459	Partial deployment	Crossed Legs		
109	BOST SCI GREENFIELD VENA CAVA	12/11/03	JWL	503580	Partial deployment	Not = 503459		
110	BOST SCI GREENFIELD FILTER	12/09/03	JWL	501625	Failed Deployment			
111	BOST SCI GREENFIELD FILTER	12/04/03	JWL	501392	Partial Deployment			
112	BOST SCI GREENFIELD FILTER	12/04/03	JWL	501392	Misplaced deployment	2 filters in report		
113	BOST SCI GREENFIELD 12FR SS	12/03/03	BPV	501755	Partial Deployment	2 filters in report		
114	BOST SCI IVC FILTER	11/18/03	BPV	196538	Filter Fracture	Filter Fracture		
115	BOST SCI TITANIUM GREENFIELD	11/12/03	BPV	496057	Misplaced Deployment			
116	BOST SCI GREENFIELD 12FR SS V	11/12/03	JWL	494850	Filter embolization	Migration (SVC)		
117	BOST SCI GREENFIELD FILTER	10/27/03	BPV	497183	Misplaced Deployment	Lot # M001505010		
118	BOST SCI GREENFIELD 12FR SS	10/08/03	BPV	489386	Other	Guidewire Entrapment		
119	BOST SCI TITANIUM GREENFIELD	10/01/03	BPV	487086	Partial Deployment			
120	BOST SCI GREENFIELD 12FR SS	09/24/03	BPV	486210	Partial Deployment			
121	BOST SCI GREENFIELD 12FR SS V	09/18/03	JWL	485644	Filter embolization	Migration (RA) / Failed Deployment		
122	BOST SCI GREENFIELD 12FR SS V	09/18/03	JWL	485728	= 485644			
123	BOST SCI GREENFIELD FILTER	08/06/03	JWL	476087	Filter embolization	Migration (Heart)		
124	BOST SCI GREENFIELD FILTER	08/06/03	JWL	476087	Failed deployment	Migration (Heart)		
125	BOST SCI GREENFIELD 12FR SS	08/01/03	BPV	476017	Misplaced Deployment			
126	BOST SCI TITANIUM GREENFIELD	07/28/03	BPV	474997	Failed Deployment	Stuck in Sheath		
127	BOST SCI TITANIUM GREENFIELD	07/23/03	BPV	474037	Partial Deployment			
128	BOST SCI GREENFIELD 12FR SS V	07/23/03	JWL	473973	Partial deployment			
129	BOST SCI TITANIUM GREENFIELD	07/21/03	BPV	473510	Filter Fracture	Filter Fracture		
130	BOST SCI TITANIUM GREENFIELD	06/24/03	JWL	468420	Misplaced Deployment			
131	BOST SCI GREENFIELD 12FR SS V	06/24/03	BPV	468242	Misplaced Deployment	Deployed Upside-down		
132	BOST SCI GREENFIELD FILTER	06/04/03	JWL	468679	Partial deployment	Migration (PA)/Partial Deployment, Lot # 50-301 (M001503010)		
133	BOST SCI GREENFIELD FILTER	06/04/03	JWL	468679	Filter embolization	Migration (PA)/Partial Deployment, Lot # 50-301 (M001503010)		
134	BOST SCI TITANIUM GREENFIELD	05/22/03	BPV	461796	Partial Deployment			
135	BOST SCI TITANIUM GREENFIELD	05/20/03	JWL	465562	Partial Deployment			
136	BOST SCI TITANIUM GREENFIELD	05/20/03	BPV	461532	Partial Deployment	2 filters reported, both failed		
137	BOST SCI TITANIUM GREENFIELD	05/20/03	BPV	461532	Partial Deployment	2 filters reported, both failed		
138	BOST SCI TITANIUM GREENFIELD	05/20/03	JWL	461540	= 461532			
139	BOST SCI TITANIUM GREENFIELD	05/19/03	JWL	461154	Partial deployment	Migration (PA)/Partial Deployment		
140	BOST SCI TITANIUM GREENFIELD	05/19/03	JWL	461154	Filter embolization	Migration (PA)/Partial Deployment		
141	BOST SCI TITANIUM GREENFIELD	05/15/03	JWL	460859	Partial deployment			
142	BOST SCI TITANIUM GREENFIELD	05/15/03	BPV	460851	Partial Deployment			
143	BOST SCI TITANIUM GREENFIELD FILTER	05/15/03	BPV	465690	Partial Deployment			
144	BOST SCI TITANIUM GREENFIELD	05/15/03	JWL	460856	Misplaced deployment			
145	BOST SCI GREENFIELD 12FR SS V	05/15/03	JWL	465722	Failed deployment	Unrelated patient death	Death	
146	BOST SCI TITANIUM GREENFIELD	05/01/03	JWL	457605	Partial deployment			
147	BOST SCI TITANIUM GREENFIELD	04/14/03	BPV	454257	Failed deployment			
148	BOST SCI TITANIUM GREENFIELD	04/02/03	BPV	456998	Partial Deployment			
149	BOST SCI GREENFIELD FILTER	03/14/03	JWL	448053	Caval perforation	1985 procedure		
150	BOST SCI TITANIUM GREENFIELD	03/07/03	BPV	447589	Failed Deployment	Stuck in Sheath		
151	BOST SCI TITANIUM GREENFIELD	02/04/03	BPV	440955	Partial Deployment			

A	B	C	D	E	F	G	H	I
152	BOST SCI GREENFIELD FILTER	01/28/03	BPV	440287	Misplaced Deployment	Lot # REF 50-300		
153	BOST SCI GREENFIELD 12FR SS V	01/24/03	BPV	439472	Partial Deployment			
154	BOST SCI TITANIUM GREENFIELD	01/21/03	BPV	439135	Misplaced Deployment			
155	BOST SCI TITANIUM GREENFIELD	01/10/03	JWL	436925	Migration	Migration (IVC)/Partial Deployment		
156	BOST SCI TITANIUM GREENFIELD	01/10/03	JWL	436925	Migration	Migration (IVC)/Partial Deployment		
157	BOST SCI GREENFIELD FILTER	01/03/03	BPV	436539	Misplaced Deployment	Lot # 50-300 (M001503000)		
158	BOST SCI GREENFIELD FILTER	12/30/02	BPV	435844	Partial Deployment	Deployed Upside-down		
159	BOST SCI TITANIUM GREENFIELD	12/13/02	BPV	432990	Partial Deployment			
160	BOST SCI TITANIUM GREENFIELD	12/13/02	BPV	432985	Partial Deployment			
161	BOST SCI TITANIUM GREENFIELD	12/13/02	BPV	432854	= 432985			
162	BOST SCI TITANIUM GREENFIELD	12/10/02	BPV	432504	Tilted filter			
163	BOST SCI GREENFIELD 12FR SS	11/08/02	JWL	426579	Partial deployment	Migration (Rt Atrium)		
164	BOST SCI GREENFIELD 12FR SS	11/08/02	JWL	426579	Filter embolization	Migration (Rt Atrium)		
165	BOST SCI GREENFIELD 12FR SS	10/03/02	BPV	420538	Misplaced Deployment			
166	BOST SCI GREENFIELD 12FR SS	09/25/02	BPV	419024	Partial Deployment			
167	BOST SCI GREENFIELD 12FR SS	09/16/02	BPV	417286	Partial Deployment			
168	BOST SCI GREENFIELD FILTER	08/19/02	BPV	413651	Partial Deployment	Lot # 4471005		
169	BOST SCI GREENFIELD 12FR SS	07/18/02	JWL	406331	Filter embolization	Migration (RA)		
170	BOST SCI TITANIUM GREENFIELD	07/02/02	BPV	403115	Other	Sheath Penetration of RV	Death	
171	BOST SCI GREENFIELD 12FR SS	06/25/02	JWL	402548	Partial Deployment	Migration (Intrahepatic)/Partial Deployment		
172	BOST SCI GREENFIELD 12FR SS	06/25/02	JWL	402548	Filter embolization	Migration (Intrahepatic)/Partial Deployment		
173	BOST SCI TITANIUM GREENFIELD	06/20/02	BPV	400981	Pulmonary Embolism	Partial Deployment/2nd filter placed	Death	
174	BOST SCI TITANIUM GREENFIELD	06/19/02	JWL	401001	Filter embolization	Migration (PA)/Partial Deployment		
175	BOST SCI TITANIUM GREENFIELD	06/19/02	JWL	400984	Filter embolization	Migration (PA)		
176	BOST SCI TITANIUM GREENFIELD	06/12/02	BPV	399412	Partial Deployment			
177	BOST SCI TITANIUM GREENFIELD	06/12/02	BPV	399383	Partial Deployment	Migration (RA)		
178	BOST SCI GREENFIELD 12FR SS	06/11/02	JWL	399256	Partial deployment	Migration (RA)		
179	BOST SCI GREENFIELD 12FR SS	06/11/02	JWL	399256	Filter embolization			
180	BOST SCI GREENFIELD 12FR SS	06/06/02	BPV	398839	Guidewire Entrapment			
181	BOST SCI TITANIUM GREENFIELD	05/22/02	BPV	396548	Misplaced Deployment			
182	BOST SCI GREENFIELD FILTER	04/18/02	BPV	390358	Partial Deployment			
183	BOST SCI TITANIUM GREENFIELD	03/29/02	BPV	386124	Partial Deployment			
184	BOST SCI GREENFIELD FILTER	03/07/02	BPV	447100	Failed Deployment	Lot # 4003871		
185	BOST SCI GREENFIELD 12FR SS	02/28/02	JWL	380010	Filter embolization	Migration (heart)		
186	BOST SCI TITANIUM GREENFIELD	02/21/02	BPV	378601	Partial Deployment			
187	BOST SCI GREENFIELD 12FR SS	02/07/02	BPV	376120	Partial Deployment			
188	BOST SCI GREENFIELD 12FR SS	01/22/02	BPV	373981	Partial Deployment			
189	BOST SCI TITANIUM GREENFIELD	01/14/02	BPV	371481	Partial Deployment			
190	BOST SCI TITANIUM GREENFIELD	12/11/01	JWL	365521	Filter embolization	Migration (femoral to SVC)		
191	BOST SCI GREENFIELD 12FR SS	12/07/01	BPV	365290	Misplaced Deployment			
192	BOST SCI TITANIUM GREENFIELD	12/05/01	BPV	364809	Partial Deployment			
193	BOST SCI GREENFIELD 12FR SS	11/19/01	BPV	362587	Partial Deployment			
194	BOST SCI GREENFIELD 12FR SS	11/08/01	BPV	360285	Partial Deployment			
195	BOST SCI TITANIUM GREENFIELD	10/22/01	BPV	358289	Partial Deployment			
196	BOST SCI GREENFIELD 12FR SS	10/22/01	BPV	358209	Partial Deployment			
197	BOST SCI GREENFIELD FILTER	10/18/01	BPV	357280	Partial Deployment	Lot # 3945902		
198	BOST SCI GREENFIELD 12FR SS	10/18/01	BPV	357289	Partial Deployment			
199	BOST SCI GREENFIELD 12FR SS	10/18/01	JWL	356485	Filter embolization	Migration (RA, then RV)		
200	BOST SCI GREENFIELD 12FR SS	10/18/01	JWL	356469	Filter embolization	Migration (PA)		
201	BOST SCI GREENFIELD 12FR SS	10/17/01	BPV	356371	Partial Deployment			
202	BOST SCI GREENFIELD 12FR SS	10/02/01	BPV	354459	Partial Deployment			
203	BOST SCI TITANIUM GREENFIELD	09/17/01	BPV	352539	Partial Deployment			

	A	B	C	D	E	F	G	H	I
204	BOST SCI	TITANIUM GREENFIELD	09/06/01	BPV	350588	Partial Deployment			
205	BOST SCI	GREENFIELD FILTER	08/24/01	BPV	349045	Pulmonary Embolism	Failed Deployment		
206	BOST SCI	GREENFIELD FILTER	08/24/01	BPV	349045	Failed Deployment	Failed Deployment	Death	
207	BOST SCI	GREENFIELD 12FR SS	08/16/01	BPV	347797	Partial Deployment			
208	BOST SCI	GREENFIELD 12FR SS	08/16/01	BPV	347784	Partial Deployment			
209	BOST SCI	GREENFIELD 12FR SS	08/15/01	BPV	347556	Partial Deployment			
210	BOST SCI	GREENFIELD 12FR SS	08/09/01	BPV	346828	Misplaced Deployment			
211	BOST SCI	TITANIUM GREENFIELD	07/25/01	BPV	343949	Partial Deployment			
212	BOST SCI	GREENFIELD 12FR SS	07/25/01	BPV	343617	Partial Deployment			
213	BOST SCI	TITANIUM GREENFIELD	07/25/01	BPV	343960	Misplaced Deployment	right atrium		
214	BOST SCI	GREENFIELD 12FR SS	07/25/01	BPV	343617	Filter embolization			
215	BOST SCI	GREENFIELD 12FR SS	07/16/01	BPV	342639	Misplaced Deployment			
216	BOST SCI	GREENFIELD 12FR SS	07/11/01	BPV	343143	Misplaced Deployment			
217	BOST SCI	GREENFIELD 12FR SS	07/09/01	BPV	341615	Misplaced Deployment			
218	BOST SCI	GREENFIELD 12FR SS	06/25/01	BPV	339558	Partial Deployment			
219	BOST SCI	GREENFIELD 12FR SS	06/22/01	BPV	339250	Partial Deployment			
220	BOST SCI	TITANIUM GREENFIELD	06/22/01	BPV	338684	Misplaced Deployment	Deployed Upside-down		
221	BOST SCI	GREENFIELD 12FR SS	06/22/01	BPV	339221	Misplaced Deployment			
222	BOST SCI	GREENFIELD 12FR SS	06/13/01	BPV	337789	Partial Deployment			
223	BOST SCI	GREENFIELD 12FR SS	06/06/01	JWL	338419	Other	Guidewire pulled filter into iliac		
224	BOST SCI	TITANIUM GREENFIELD	06/06/01	JWL	336293	Filter embolization	Migration (PA)		
225	BOST SCI	GREENFIELD FILTER	05/17/01	BPV	333639	Misplaced Deployment			
226	BOST SCI	GREENFIELD 12FR SS	04/24/01	BPV	329109	Partial Deployment	Filter Fracture		
227	BOST SCI	TITANIUM GREENFIELD	04/20/01	BPV	328313	Filter fracture			
228	BOST SCI	GREENFIELD 12FR SS	04/10/01	BPV	325919	Caval Perforation			
229	BOST SCI	GREENFIELD 12FR SS	04/06/01	JWL	324695	Filter embolization	Migration (Rt atrium)/GW entrapment	Death	
230	BOST SCI	TITANIUM GREENFIELD	03/16/01	BPV	321369	Partial Deployment			
231	BOST SCI	GREENFIELD 12FR SS	03/09/01	BPV	320035	Caval Perforation			
232	BOST SCI	TITANIUM GREENFIELD	02/28/01	JWL	319266	Partial deployment			
233	BOST SCI	GREENFIELD 12FR SS	02/20/01	BPV	317528	Partial Deployment			
234	BOST SCI	GREENFIELD 12FR SS	02/14/01	BPV	316104	Partial Deployment			
235	BOST SCI	GREENFIELD 12FR SS	02/09/01	JWL	315274	Partial deployment	Migration (Cephalad)		
236	BOST SCI	GREENFIELD FILTER	02/09/01	JWL	315264	Partial deployment	Migration (Cephalad)		
237	BOST SCI	GREENFIELD 12FR SS	02/09/01	JWL	315274	Migration	Migration (Cephalad)		
238	BOST SCI	GREENFIELD FILTER	02/09/01	JWL	315264	Migration	Migration (Cephalad)		
239	BOST SCI	GREENFIELD 12FR SS	02/08/01	BPV	315122	Guidewire Entrapment			
240	BOST SCI	GREENFIELD FILTER	02/02/01	BPV	314449	Caval Perforation		Death	
241	BOST SCI	GREENFIELD 12FR SS	01/25/01	BPV	313979	Partial Deployment			
242	BOST SCI	GREENFIELD 12FR SS	01/17/01	BPV	312972	Partial Deployment			
243	BOST SCI	TITANIUM GREENFIELD	01/16/01	BPV	312588	Partial Deployment			
244	BOST SCI	GREENFIELD 12FR SS	01/10/01	BPV	312130	Partial Deployment			
245	BOST SCI	GREENFIELD 12FR SS	01/10/01	BPV	312124	Partial Deployment			
246	BOST SCI	GREENFIELD FILTER	01/05/01	JWL	311496	Filter embolization	Migration (Rt Atrium)		
247	BOST SCI	GREENFIELD 12FR SS	01/04/01	JWL	311046	Partial deployment	Migration (Heart)		
248	BOST SCI	GREENFIELD 12FR SS	01/04/01	JWL	311046	Filter embolization	Migration (Heart)		
249	BOST SCI	TITANIUM GREENFIELD	12/20/00	BPV	309973	Partial Deployment			
250	BOST SCI	GREENFIELD 12FR SS	12/19/00	BPV	309310	Guidewire Entrapment			
251	BOST SCI	TITANIUM GREENFIELD	12/08/00	BPV	308044	Partial Deployment			
252	BOST SCI	GREENFIELD 12FR SS	12/07/00	BPV	307740	Partial Deployment			
253	BOST SCI	GREENFIELD 12FR SS	12/07/00	BPV	307729	Filter fracture	Filter Fracture		
254	BOST SCI	GREENFIELD 12FR SS	12/05/00	BPV	307673	Partial Deployment			
255	BOST SCI	GREENFIELD 12FR SS	11/27/00	JWL	306489	Migration	Migration (Rt Atrium)/Guidewire Entrapment		

	A	B	C	D	E	F	G	H	I
256	BOST SCI	GREENFIELD FILTER	11/27/00	JWL	306487	Migration	Migration (Iliac)		
257	BOST SCI	GREENFIELD FILTER	11/17/00	BPV	305247	Misplaced Deployment	Deployed Upside-down		
258	BOST SCI	GREENFIELD 12FR SS	11/10/00	BPV	304539	Partial Deployment			
259	BOST SCI	GREENFIELD 12FR SS	11/01/00	JWL	302808	Filter embolization	Migration (PA)/Partial Deployment		
260	BOST SCI	GREENFIELD 12FR SS	10/31/00	BPV	302837	Partial Deployment			
261	BOST SCI	GREENFIELD 24FR SS V	10/27/00	JWL	301962	Filter fracture	Filter Fracture		
262	BOST SCI	GREENFIELD FILTER	10/24/00	BPV	301521	Caval Perforation	Central Line	Death	
263	BOST SCI	GREENFIELD 12FR SS	10/03/00	BPV	299224	Guidewire Entrapment			
264	BOST SCI	GREENFIELD 12FR SS	09/27/00	BPV	298479	Partial Deployment			
265	BOST SCI	TITANIUM GREENFIELD	09/27/00	BPV	298485	Misplaced Deployment			
266	BOST SCI	TITANIUM GREENFIELD	09/19/00	BPV	296859	Partial Deployment			
267	BOST SCI	TITANIUM GREENFIELD	09/19/00	BPV	296855	Partial Deployment			
268	BOST SCI	TITANIUM GREENFIELD	08/12/00	BPV	295758	Filter fracture	Filter Fracture		
269	BOST SCI	TITANIUM GREENFIELD	09/05/00	BPV	294486	Partial Deployment			
270	BOST SCI	GREENFIELD 12FR SS	09/05/00	BPV	294217	Partial Deployment			
271	BOST SCI	GREENFIELD 12FR SS	09/05/00	BPV	294495	Misplaced Deployment	Deployed Upside-down		
272	BOST SCI	TITANIUM GREENFIELD	09/01/00	BPV	294503	Caval Perforation			
273	BOST SCI	GREENFIELD 12FR SS	08/23/00	BPV	292346	Guidewire Entrapment			
274	BOST SCI	GREENFIELD 12FR SS	08/14/00	BPV	290202	Partial Deployment			
275	BOST SCI	TITANIUM GREENFIELD	08/02/00	BPV	288431	Misplaced Deployment			
276	BOST SCI	GREENFIELD 12FR SS	08/02/00	BPV	288163	Failed Deployment			
277	BOST SCI	GREENFIELD FILTER	07/26/00	BPV	287223	Filter fracture			
278	BOST SCI	GREENFIELD 12FR SS	07/18/00	BPV	286578	Partial Deployment			
279	BOST SCI	GREENFIELD 12FR SS	07/18/00	BPV	286574	Partial Deployment			
280	BOST SCI	GREENFIELD 12FR SS	07/18/00	BPV	286451	Misplaced Deployment			
281	BOST SCI	GREENFIELD 12FR SS	07/18/00	BPV	286446	Misplaced Deployment			
282	BOST SCI	TITANIUM GREENFIELD	06/21/00	BPV	283348	Partial Deployment			
283	BOST SCI	TITANIUM GREENFIELD	06/21/00	BPV	282884	Partial Deployment			
284	BOST SCI	GREENFIELD 12FR SS	06/21/00	BPV	283341	Partial Deployment			
285	BOST SCI	GREENFIELD 12FR SS	06/21/00	BPV	283262	Guidewire Entrapment			
286	BOST SCI	GREENFIELD FILTER	06/12/00	BPV	282506	Partial Deployment			
287	BOST SCI	GREENFIELD 12FR SS	06/12/00	BPV	281960	Failed Deployment			
288	BOST SCI	GREENFIELD 12FR SS	06/07/00	BPV	281518	Misplaced Deployment			
289	BOST SCI	TITANIUM GREENFIELD	06/06/00	JWL	280475	Filter embolization	Migration (PA)		
290	BOST SCI	TITANIUM GREENFIELD	05/30/00	BPV	280085	Pulmonary Embolism	Caval Thrombosis	Death	
291	BOST SCI	TITANIUM GREENFIELD	05/30/00	BPV	280454	Partial Deployment			
292	BOST SCI	GREENFIELD 12FR SS	05/30/00	BPV	280457	Partial Deployment			
293	BOST SCI	GREENFIELD 12FR SS	05/30/00	BPV	280460	Misplaced Deployment			
294	BOST SCI	GREENFIELD 12FR SS	05/30/00	BPV	280458	Misplaced Deployment			
295	BOST SCI	GREENFIELD 12FR SS	05/30/00	BPV	280457	Migration	into suprarenal IVC		
296	BOST SCI	TITANIUM GREENFIELD	05/30/00	JWL	280287	Filter embolization	PA		
297	BOST SCI	TITANIUM GREENFIELD	05/30/00	JWL	280281	= 280287	PA		
298	BOST SCI	GREENFIELD 12FR SS	05/23/00	BPV	279283	Partial Deployment			
299	BOST SCI	GREENFIELD 12FR SS	05/18/00	BPV	278823	Partial Deployment			
300	BOST SCI	TITANIUM GREENFIELD	05/16/00	JWL	278192	Partial deployment	Right atrium		
301	BOST SCI	TITANIUM GREENFIELD	05/16/00	JWL	277714	Filter embolization			
302	BOST SCI	TITANIUM GREENFIELD	05/16/00	BPV	277510	Caval Perforation			
303	BOST SCI	TITANIUM GREENFIELD	05/15/00	BPV	278198	Misplaced Deployment	Guidewire Fracture	Death	
304	BOST SCI	TITANIUM GREENFIELD	05/15/00	BPV	278203	Failed Deployment			
305	BOST SCI	TITANIUM GREENFIELD	04/25/00	BPV	275192	Partial Deployment			
306	BOST SCI	GREENFIELD FILTER	04/25/00	JWL	275896	Filter embolization	Migration (Rt Atrium)		
307	BOST SCI	GREENFIELD FILTER	04/13/00	JWL	273523	Other		Death	

A	B	C	D	E	F	G	H	I
308	BOST SCI GREENFIELD FILTER	04/05/00	BPV	272392	Pulmonary Embolism			
309	BOST SCI GREENFIELD 12FR SS	03/30/00	BPV	271789	Partial Deployment		Death	
310	BOST SCI GREENFIELD 12FR SS	03/23/00	BPV	270582	Guidewire entrapment			
311	BOST SCI GREENFIELD 12FR SS	03/23/00	BPV	270582	Caval perforation			
312	BOST SCI GREENFIELD 12FR SS	03/13/00	BPV	269355	Partial Deployment			
313	BOST SCI GREENFIELD FILTER	02/29/00	JWL	266358	Misplaced Deployment			
314	BOST SCI TITANIUM GREENFIELD	02/24/00	BPV	264706	Partial Deployment			
315	BOST SCI TITANIUM GREENFIELD	02/15/00	JWL	262672	Filter embolization	Migration (Rt Atrium)/Partial Deployment		
316	BOST SCI GREENFIELD 12FR SS	02/08/00	JWL	261703	Filter embolization			
317	BOST SCI GREENFIELD 12FR SS	02/07/00	BPV	261685	Guidewire entrapment			
318	BOST SCI GREENFIELD 12FR SS	02/03/00	BPV	261346	Partial Deployment			
319	BOST SCI TITANIUM GREENFIELD	01/24/00	BPV	260104	Partial Deployment			
320	BOST SCI GREENFIELD 12FR SS	01/24/00	BPV	259996	Partial Deployment			
321	BOST SCI GREENFIELD 12FR SS	01/20/00	BPV	260088	Misplaced deployment	Right atrium		
322	BOST SCI GREENFIELD 12FR SS	01/20/00	BPV	260088	Filter embolization	Right atrium		
323	BOST SCI TITANIUM GREENFIELD	01/19/00	BPV	259451	Partial Deployment			
324	BOST SCI GREENFIELD 12FR SS	01/13/00	BPV	258915	Partial Deployment			
325	BOST SCI GREENFIELD FILTER	01/11/00	BPV	258378	Partial Deployment			
326	BOST SCI GREENFIELD FILTER	01/11/00	BPV	258378	Filter embolization	right atrium		
327	BOST SCI TITANIUM GREENFIELD	01/06/00	BPV	258202	Partial Deployment			
328	BOST SCI GREENFIELD 12FR SS	01/06/00	BPV	258192	Guidewire entrapment			
329	BOST SCI GREENFIELD 12FR SS	01/06/00	BPV	257774	Guidewire entrapment			
330	BOST SCI GREENFIELD 12FR SS	01/05/00	JWL	257468	Guidewire entrapment			

[illegible]

	A	B	C	D	E	F	G	H	I	J	K
	Death	Migration	Filter Embolization	Caval Perforation	Caval Thrombosis	Pulmonary Embolism	Filter Fracture	Guidewire Entrapment	Misplaced Deployment	Partial Deployment	Failed Deployment
1	7	6	8	10	2	2	0	0	5	17	13
2											
3											
4											
5											
6	Company	Device	Date	Source	FDA #	Failure	Notes	Death?	FilterEmb death?	2	
7	COOK	GUNTHER TULIP VENA C	9/23/2005	KMJ	636844	Failed deployment	at implant				
8	COOK	GUNTHER TULIP VENA C	9/9/2005	KMJ	634010	Migration					
9	COOK	GUNTHER TULIP	9/9/2005	KMJ	634005	Failed deployment					
10	COOK	GUNTHER TULIP VENA C	9/7/2005	KMJ	633458	Failed deployment	vessel pierced?				
11	COOK	GUNTHER TULIP VENA C	7/27/2005	KMJ	622800	Other	bled profusely from access site				
12	COOK	GUNTHER TULIP VENA C	7/27/2005	KMJ	622818	Partial deployment					
13	COOK, INC.	GUNTHER TULIP VENA C	6/15/2005	KMJ	613925	Detachment of Component(s)	Guidewire & sheath in heart	Death			
14	COOK, INC.	GUNTHER TULIP VENA C	5/27/2005	KMJ	608107	Caval Perforation					
15	COOK, INC.	GUNTHER TULIP VENA C	4/27/2005	KMJ	597905	Partial deployment					
16	COOK, INC.	GUNTHER TULIP VENA C	4/8/2005	KMJ	589113	Partial deployment					
17	COOK, INC.	GUNTHER TULIP VENA C	4/5/2005	KMJ	587808	Pulmonary embolism	?				
18	COOK, INC.	GUNTHER TULIP VENA C	4/12/2005	KMJ	587043	Partial deployment	"opposed itself to the cava wall"				
19	COOK, INC.	GUNTHER TULIP VENA C	3/25/2005	KMJ	584933	Tilted Filter	Autopsy results pending	Death			
20	COOK, INC.	GUNTHER TULIP VENA C	3/4/2005	KMJ	578055	Misplaced Deployment					
21	COOK, INC.	GUNTHER TULIP VENA C	3/4/2005	KMJ	578055	Caval Perforation					
22	COOK, INC.	UNKNOWN BRAND NAME	2/25/2005	KMJ	576360	Other	Unknown injury				
23	COOK, INC.	GUNTHER TULIP VENA C	2/24/2005	KMJ	575851	Partial deployment	Legs did not fully open				
24	COOK, INC.	GUNTHER TULIP VENA C	2/18/2005	KMJ	574713	Partial deployment	Legs did not fully open				
25	COOK, INC.	GUNTHER TULIP VENA C	2/18/2005	KMJ	574732	Partial deployment	Filter opened asymmetrically				
26	COOK, INC.	GUNTHER TULIP VENA C	2/18/2005	KMJ	574702	Partial deployment	Legs did not fully open				
27	COOK, INC.	GUNTHER TULIP VENA C	2/4/2005	KMJ	570330	Partial deployment	Legs did not fully open				
28	COOK, INC.	GUNTHER TULIP VENA C	1/27/2005	KMJ	542855	Failed deployment		Death			
29	COOK, INC.	GUNTHER TULIP VENA C	10/21/04	BPV	550370	Caval Thrombosis					
30	COOK, INC.	GUNTHER TULIP VENA C	10/15/04	BPV	548881	Failed deployment					
31	COOK, INC.	GUNTHER TULIP VENA C	10/08/04	BPV	547688	Caval Perforation					
32	COOK, INC.	GUNTHER TULIP VENA C	8/26/2004	BPV	540894	Partial deployment	crossed legs				
33	COOK, INC.	GUNTHER TULIP VENA C	7/23/2004	BPV	535670	Caval Thrombosis					
34	COOK, INC.	GUNTHER TULIP VENA C	6/3/2004	BPV	534699	Caval Perforation					
35	COOK, INC.	GUNTHER TULIP	05/07/04	BPV	525493	Other	Defective Part				
36	COOK, INC.	GUNTHER TULIP	05/07/04	BPV	524016	Other	Broken Snare				
37	COOK, INC.	GUNTHER TULIP VENA C	5/7/2004	BPV	524016	Other	retrieval failure				
38	COOK, INC.	GUNTHER TULIP	04/22/04	BPV	521847	Tilted Filter					
39	COOK, INC.	GUNTHER TULIP	04/14/04	BPV	520682	Failed deployment					
40	COOK, INC.	GUNTHER TULIP	04/08/04	BPV	519818	Migration					
41	COOK, INC.	GUNTHER TULIP	04/08/04	BPV	519818	Caval Perforation					
42	COOK, INC.	GUNTHER TULIP	03/22/04	BPV	517127	Migration	Migration (Suprarenal)				
43	COOK, INC.	GUNTHER TULIP	03/19/04	BPV	516823	Caval Perforation					
44	COOK, INC.	GUNTHER TULIP VENA C	02/11/04	JWL	510942	Failed deployment					
45	COOK, INC.	GUNTHER TULIP	02/11/04	BPV	510932	Failed deployment	Filter Did Not Open				
46	COOK, INC.	GUNTHER TULIP VENA C	12/29/03	JWL	503753	Filter embolization	Rt Atrium				
47	COOK, INC.	GUNTHER TULIP VENA C	11/20/03	BPV	497932	Partial Deployment					
48	COOK, INC.	GUNTHER TULIP VENA C	10/28/03	BPV	491788	Partial Deployment					
49	COOK, INC.	GUNTHER TULIP VENA C	10/14/03	BPV	488736	Failed Deployment					

[illegible]

	A	B	C	D	E	F	G	H	I	J
	Death	Migration	Filter Embolization	Caval Perforation	Caval Thrombosis	Pulmonary Embolism	Filter Fracture	Guidewire Entrapment	Misplaced Deployment	Partial Deployment
1	29	10	19	19	62	8	18	0	0	15
2										
3										
4										
5										
6	Company	Device	Date	Source	FDA #	Failure	Notes	Death?	Filter Emb death?	9
7	CORDIS EUROPA, N.V.	TRAPEASE PERMANENT V	9/8/2005	KMJ	634320	Failed Deployment				
8	CORDIS EUROPA, N.V.	TRAPEASE PERMANENT V	9/7/2005	KMJ	634304	Filter Fracture				
9	CORDIS EUROPA, N.V.	TRAPEASE PERMANENT V	8/30/2005	KMJ	633190	Caval Thrombosis				
10	CORDIS EUROPA, N.V.	TRAPEASE PERMANENT V	8/25/2005	KMJ	632044	Failed Deployment				
11	CORDIS EUROPA, N.V.	TRAPEASE PERMANENT V	8/22/2005	KMJ	629334	Caval Thrombosis				
12	CORDIS EUROPA, N.V.	TRAPEASE PERMANENT V	8/12/2005	KMJ	627058	Pulmonary embolism		Death		
13	CORDIS EUROPA, N.V.	TRAPEASE PERMANENT V	8/5/2005	KMJ	624909	Caval Thrombosis				
14	CORDIS EUROPA, N.V.	TRAPEASE PERMANENT V	7/18/2005	KMJ	622017	Pulmonary embolism		Death		
15	CORDIS EUROPA, N.V.	TRAPEASE PERMANENT V	7/18/2005	KMJ	622058	Failed Deployment				
16	CORDIS EUROPA, N.V.	TRAPEASE PERMANENT F	7/14/2005	KMJ	621741	Migration	to diaphragm			
17	CORDIS EUROPA, N.V.	TRAPEASE PERMANENT V	6/23/2005	KMJ	617473	Filter embolization				
18	CORDIS EUROPA, N.V.	TRAPEASE PERMANENT V	6/10/2005	KMJ	613267	Filter embolization	Morbidly obese pt	Death		
19	CORDIS EUROPA, N.V.	TRAPEASE UNK FILTER	6/2/2005	KMJ	611082	Other	laceration 2cm distal to filter	Death		
20	CORDIS EUROPA, N.V.	TRAPEASE PERMANENT V	4/29/2005	KMJ	598106	Filter embolization	VC fully thrombosed	Death	1	
21	CORDIS EUROPA, N.V.	TRAPEASE PERMANENT V	4/27/2005	KMJ	597684	Failed Deployment				
22	CORDIS EUROPA, N.V.	TRAPEASE PERMANENT V	4/22/2005	KMJ	595176	Other	Retroperitoneal bleed			
23	CORDIS EUROPA, N.V.	TRAPEASE PERMANENT V	4/8/2005	KMJ	590441	Migration	Suprarenal			
24	CORDIS EUROPA, N.V.	TRAPEASE UNK FILTER	3/21/2005	KMJ	584046	Caval Thrombosis				
25	CORDIS EUROPA, N.V.	TRAPEASE PERMANENT V	3/18/2005	KMJ	583561	Failed Deployment				
26	CORDIS EUROPA, N.V.	TRAPEASE PERMANENT V	3/16/2005	KMJ	582555	Caval Perforation		Death		
27	CORDIS EUROPA, N.V.	TRAPEASE PERMANENT V	3/10/2005	KMJ	581289	Caval Thrombosis				
28	CORDIS EUROPA, N.V.	TRAPEASE UNK FILTER	3/7/2005	KMJ	579168	Filter embolization	Heart			
29	CORDIS EUROPA, N.V.	TRAPEASE PERMANENT V	3/3/2005	KMJ	581668	Filter Fracture				
30	CORDIS EUROPA, N.V.	TRAPEASE UNKNOWN FIL	2/28/2005	KMJ	577727	Filter embolization	Pulmonary artery	Death	1	
31	CORDIS EUROPA, N.V.	TRAPEASE PERMANENT V	2/25/2005	KMJ	580543	Filter Fracture				
32	CORDIS EUROPA, N.V.	TRAPEASE PERMANENT VE	2/25/2005	KMJ	576690	Filter Fracture				
33	CORDIS EUROPA, N.V.	TRAPEASE UNK FILTER	2/24/2005	KMJ	576373	Caval Perforation	fracture punctured aorta	Death		
34	CORDIS EUROPA, N.V.	TRAPEASE UNK FILTER	2/24/2005	KMJ	576373	Filter Fracture	fracture punctured aorta	Death		
35	CORDIS EUROPA, N.V.	TRAPEASE PERMANENT F	2/22/2005	KMJ	574994	Caval Perforation	Perforation + Migration	Death		
36	CORDIS EUROPA, N.V.	TRAPEASE PERMANENT F	2/22/2005	KMJ	574994	Migration	Perforation + Migration	Death		
37	CORDIS EUROPA, N.V.	TRAPEASE PERMANENT V	2/22/2005	KMJ	577648	Filter Fracture				
38	CORDIS EUROPA, N.V.	TRAPEASE VENA CAVA F	2/15/2005	KMJ	553138	Filter embolization	autopsy pending	Death	1	
39	CORDIS EUROPA, N.V.	TRAPEASE PERMANENT V	2/11/2005	KMJ	573024	Failed Deployment				
40	CORDIS EUROPA, N.V.	TRAPEASE PERMANENT V	2/11/2005	KMJ	576426	Failed Deployment				
41	CORDIS EUROPA, N.V.	TRAPEASE PERMANENT V	2/11/2005	KMJ	572540	Pulmonary embolism	PE & Fracture (same pt)			
42	CORDIS EUROPA, N.V.	TRAPEASE PERMANENT V	1/28/2005	KMJ	581062	Filter Fracture	PE & Fracture (same pt)			
43	CORDIS EUROPA, N.V.	TRAPEASE PERMANENT V	1/21/2005	KMJ	583297	Failed Deployment				
44	CORDIS EUROPA, N.V.	TRAPEASE PERMANENT V	1/5/2005	BPV	567329	Partial Deployment				
45	CORDIS CORPORATION	TRAPEASE PERMANENT V	11/22/04	BPV	556636	Filter embolization	Tricuspid Valve			
46	CORDIS CORPORATION	TRAPEASE PERMANENT V	11/17/04	BPV	555663	Other				
47	CORDIS CORPORATION	TRAPEASE IVC FILTER	11/01/04	BPV	558588	Filter Fracture	punctured aorta	Death		
48	CORDIS CORPORATION	TRAPEASE FILTER	10/28/04	BPV	561521	Caval Thrombosis				
49	CORDIS CORPORATION	TRAPEASE PERMANENT V	8/12/2004	BPV	538721	Caval Thrombosis				
50	CORDIS CORPORATION	TRAPEASE PERMANENT V	8/12/2004	BPV	538729	Caval Thrombosis				
51	CORDIS CORPORATION	TRAPEASE PERMANENT V	8/12/2004	BPV	538833	Caval Thrombosis				
52	CORDIS CORPORATION	TRAPEASE FILTER	7/16/2004	BPV	535731	Caval Perforation		Death		
53	CORDIS CORPORATION	TRAPEASE FILTER	7/16/2004	BPV	535731	Caval thrombosis				
54	CORDIS CORPORATION	TRAPEASE PERMANENT V	7/13/2004	BPV	534357	Failed Deployment				
55	CORDIS CORPORATION	TRAPEASE PERMANENT V	7/13/2004	BPV	534359	Filter embolization	heart PE	Death	1	
56	CORDIS CORPORATION	TRAP-EASE	7/1/2004	BPV	533876	Other		Death		

A	B	C	D	E	F	G	H	I	J
Company	Device	Date	Source	FDA #	Failure	Notes	Death?	FilterEmb death?	
6									9
57	CORDIS CORPORATION	TRAPEASE PERMANENT V	6/28/2004	BPV	532518	Filter embolization			
58	CORDIS CORPORATION	TRAPEASE PERMANENT V	06/18/04	BPV	530718	Caval Perforation			
59	CORDIS CORPORATION	TRAPEASE VENA CAVA F	4/30/2004	BPV	525715	Filter embolization			
60	CORDIS CORPORATION	TRAPEASE VENA CAVA F	4/30/2004	BPV	525715	Failed deployment			
61	CORDIS CORPORATION	TRAPEASE PERMANENT V	04/28/04	BPV	522537	Filter Fracture			
62	CORDIS CORPORATION	TRAPEASE PERMANENT V	04/28/04	BPV	522537	Pulmonary embolism			
63	CORDIS CORPORATION	TRAPEASE PERMANENT V	03/01/04	BPV	514509	Caval Perforation			
64	CORDIS CORPORATION	TRAPEASE PERMANENT V	02/23/04	BPV	512770	Failed Deployment			
65	CORDIS CORPORATION	TRAPEASE PERMANENT V	2/4/2004	BPV	525165	Filter Fracture			
66	CORDIS EUROPA, N.V.	TRAPEASE PERMANENT V	2/2/2004	JWL	509799	= 510344			
67	CORDIS CORPORATION	TRAPEASE PERMANENT V	02/02/04	BPV	510344	Caval Thrombosis			
68	CORDIS CORPORATION	TRAPEASE PERMANENT V	01/20/04	BPV	513000	= 513006			
69	CORDIS CORPORATION	TRAPEASE PERMANENT V	1/20/2004	BPV	513006	Other			
70	CORDIS CORPORATION	TRAPEASE PERMANENT V	01/09/04	BPV	505640	Filter embolization			
71	CORDIS CORPORATION	TRAPEASE PERMANENT V	12/19/03	BPV	501966	Other			
72	CORDIS CORPORATION	TRAPEASE PERMANENT V	12/19/03	BPV	501994	Migration			
73	CORDIS CORPORATION	TRAPEASE PERMANENT V	12/05/03	BPV	502004	Filter Fracture			
74	CORDIS CORPORATION	TRAPEASE PERMANENT V	11/20/03	BPV	496583	Filter Fracture			
75	CORDIS CORPORATION	TRAPEASE PERMANENT V	11/14/03	BPV	495497	Filter embolization			
76	CORDIS CORPORATION	TRAPEASE PERMANENT V	11/11/03	BPV	495600	Filter embolization			
77	CORDIS CORPORATION	TRAPEASE PERMANENT V	11/05/03	BPV	493714	Caval Thrombosis			
78	CORDIS CORPORATION	TRAPEASE PERMANENT V	10/31/03	BPV	493717	Caval Thrombosis			
79	CORDIS CORPORATION	TRAPEASE PERMANENT V	10/31/03	BPV	494160	Filter Fracture			
80	CORDIS CORPORATION	TRAPEASE PERMANENT V	10/29/03	BPV	492611	Caval Thrombosis			
81	CORDIS CORPORATION	TRAPEASE PERMANENT V	10/27/03	BPV	491452	Caval Perforation			
82	CORDIS CORPORATION	TRAPEASE PERMANENT V	10/27/03	BPV	492352	Caval Perforation			
83	CORDIS CORPORATION	TRAPEASE PERMANENT V	10/08/03	BPV	489092	Filter embolization			
84	CORDIS CORPORATION	TRAPEASE PERMANENT V	10/08/03	BPV	490223	Caval Perforation			
85	CORDIS CORPORATION	TRAPEASE PERMANENT V	09/15/03	BPV	487678	Caval Thrombosis			
86	CORDIS CORPORATION	TRAPEASE PERMANENT V	09/15/03	BPV	487682	Caval Thrombosis			
87	CORDIS CORPORATION	TRAPEASE PERMANENT V	09/10/03	BPV	483427	Partial Deployment			
88	CORDIS CORPORATION	TRAPEASE PERMANENT V	08/12/03	BPV	478085	Filter Fracture			
89	CORDIS CORPORATION	TRAPEASE PERMANENT V	08/07/03	BPV	477020	Filter Fracture			
90	CORDIS CORPORATION	TRAPEASE PERMANENT V	08/06/03	BPV	476182	Caval Thrombosis			
91	CORDIS CORPORATION	TRAPEASE PERMANENT V	08/01/03	BPV	475628	= 471736			
92	CORDIS CORPORATION	TRAPEASE PERMANENT V	07/23/03	BPV	473871	Partial Deployment			
93	CORDIS CORPORATION	TRAPEASE PERMANENT V	07/15/03	BPV	471736	Filter embolization			
94	CORDIS CORPORATION	TRAPEASE PERMANENT V	06/27/03	BPV	473028	= 471736			
95	CORDIS CORPORATION	TRAPEASE PERMANENT V	06/18/03	BPV	466546	Caval Perforation			
96	CORDIS CORPORATION	TRAPEASE PERMANENT V	06/17/03	BPV	466036	Partial Deployment			
97	CORDIS CORPORATION	TRAPEASE PERMANENT V	04/28/03	BPV	456208	Caval Thrombosis			
98	CORDIS CORPORATION	TRAPEASE PERMANENT V	03/20/03	BPV	449105	Other			
99	CORDIS CORPORATION	TRAPEASE PERMANENT V	02/28/03	BPV	445805	= 445807			
100	CORDIS CORPORATION	TRAPEASE PERMANENT V	02/28/03	BPV	445807	Caval Thrombosis			
101	CORDIS CORPORATION	TRAPEASE PERMANENT V	02/28/03	BPV	445967	= 445807			
102	CORDIS CORPORATION	TRAPEASE PERMANENT V	12/19/02	BPV	434341	Caval Thrombosis			
103	CORDIS CORPORATION	TRAPEASE PERMANENT V	11/26/02	BPV	429685	Caval Thrombosis			
104	CORDIS CORPORATION	TRAPEASE PERMANENT V	11/26/02	BPV	430340	Caval Thrombosis			
105	CORDIS CORPORATION	TRAPEASE PERMANENT V	10/31/02	BPV	425265	Caval Perforation			
106	CORDIS CORPORATION	TRAPEASE PERMANENT V	10/23/02	BPV	424103	Failed Deployment			
107	CORDIS CORPORATION	TRAPEASE PERMANENT V	10/23/02	BPV	424106	Other			
108	CORDIS CORPORATION	TRAPEASE PERMANENT V	10/15/02	BPV	422521	Caval Thrombosis			
109	CORDIS CORPORATION	TRAPEASE PERMANENT V	10/12/02	BPV	422969	Caval Perforation			
110	CORDIS CORPORATION	TRAPEASE PERMANENT V	10/04/02	BPV	421028	Caval Thrombosis			
111	CORDIS CORPORATION	TRAPEASE PERMANENT V	10/04/02	BPV	421045	= 421028			
112	CORDIS CORPORATION	TRAPEASE PERMANENT V	10/04/02	BPV	421059	= 421028			
113	CORDIS CORPORATION	TRAPEASE PERMANENT V	10/04/02	BPV	421132	= 421028			

	A	B	C	D	E	F	G	H	I	J
	Company	Device	Date	Source	FDA #	Failure	Notes	Death?	Filter Emb death?	
6	CORDIS CORPORATION	TRAPEASE PERMANENT V	10/04/02	BPV	421155	= 421028				9
114	CORDIS CORPORATION	TRAPEASE PERMANENT V	10/04/02	BPV	421155	Caval perforation				
115	CORDIS CORPORATION	TRAPEASE PERMANENT V	10/03/02	BPV	420438	Migration	Migration (Suprarenal)			
116	CORDIS CORPORATION	TRAPEASE PERMANENT V	10/02/02	BPV	419378	Migration	Migration (Suprarenal)			
117	CORDIS CORPORATION	TRAPEASE PERMANENT V	10/01/02	BPV	419891	Filter Fracture				
118	CORDIS CORPORATION	TRAPEASE PERMANENT V	09/02/02	BPV	414795	= 414801				
119	CORDIS CORPORATION	TRAPEASE PERMANENT V	09/02/02	BPV	414801	Caval Thrombosis		Death		
120	CORDIS CORPORATION	TRAPEASE PERMANENT V	09/02/02	BPV	414807	Caval Thrombosis				
121	CORDIS CORPORATION	TRAPEASE PERMANENT V	09/02/02	BPV	414830	Caval Thrombosis				
122	CORDIS CORPORATION	TRAPEASE PERMANENT V	08/30/02	BPV	414493	Partial Deployment				
123	CORDIS CORPORATION	TRAPEASE PERMANENT V	08/30/02	BPV	414493	Migration				
124	CORDIS CORPORATION	TRAPEASE PERMANENT V	08/30/02	BPV	412485	Pulmonary embolism				
125	CORDIS CORPORATION	TRAPEASE PERMANENT V	08/09/02	BPV	409571	Migration	Bariatric use	Death		
126	CORDIS CORPORATION	TRAPEASE PERMANENT V	07/03/02	BPV	402872	Failed deployment	Migration			
127	CORDIS CORPORATION	TRAPEASE PERMANENT V	07/01/02	BPV	403107	Partial Deployment		Death		
128	CORDIS CORPORATION	TRAPEASE PERMANENT V	06/20/02	BPV	400720	Other	Delivery sheath failure			
129	CORDIS CORPORATION	TRAPEASE PERMANENT V	06/10/02	BPV	398630	Caval Thrombosis				
130	CORDIS CORPORATION	TRAPEASE PERMANENT V	06/10/02	BPV	398778	Caval thrombosis	Inferred from MDR			
131	CORDIS CORPORATION	TRAPEASE PERMANENT V	06/03/02	BPV	397536	Caval Thrombosis				
132	CORDIS CORPORATION	TRAPEASE PERMANENT V	05/29/02	BPV	396809	Caval Perforation				
133	CORDIS CORPORATION	TRAPEASE PERMANENT V	05/29/02	BPV	396812	Caval Perforation				
134	CORDIS CORPORATION	TRAPEASE PERMANENT V	03/25/02	BPV	384621	Caval Thrombosis				
135	CORDIS CORPORATION	TRAPEASE PERMANENT V	03/19/02	BPV	382807	Caval Thrombosis				
136	CORDIS CORPORATION	TRAPEASE PERMANENT V	03/19/02	BPV	382812	Caval Thrombosis				
137	CORDIS CORPORATION	TRAPEASE PERMANENT V	03/19/02	BPV	382814	Caval Thrombosis				
138	CORDIS CORPORATION	TRAPEASE PERMANENT V	03/19/02	BPV	382828	Caval Thrombosis				
139	CORDIS CORPORATION	TRAPEASE PERMANENT V	03/19/02	BPV	382832	Caval Thrombosis				
140	CORDIS CORPORATION	TRAPEASE PERMANENT V	03/08/02	BPV	381138	Other	Letter of Concern			
141	CORDIS CORPORATION	TRAPEASE PERMANENT V	01/30/02	BPV	374525	Migration	Migration (Suprarenal)			
142	CORDIS CORPORATION	TRAPEASE PERMANENT V	01/30/02	BPV	374525	Other	Renal vein thrombosis			
143	CORDIS CORPORATION	TRAPEASE PERMANENT V	01/30/02	BPV	374525	Other				
144	CORDIS CORPORATION	TRAPEASE PERMANENT V	01/30/02	BPV	374620	Filter Fracture				
145	CORDIS CORPORATION	TRAPEASE PERMANENT V	12/04/01	BPV	363383	Caval perforation				
146	CORDIS CORPORATION	TRAPEASE PERMANENT V	11/01/01	BPV	359223	Failed Deployment	Stuck in Sheath			

A	B	C	D	E	F	G	H	I	J
Company	Device	Date	Source	FDA #	Failure	Notes	Death?	Fill/Emb death?	
6									
147	CORDIS CORPORATION	10/17/01	JWL	356045	Filter embolization	Migration (Rt Atrium)	Death	1	9
148	CORDIS CORPORATION	09/27/01	BPV	353318	Caval Thrombosis				
149	CORDIS CORPORATION	09/19/01	BPV	352239	Caval Thrombosis				
150	CORDIS CORPORATION	09/10/01	BPV	351038	Partial Deployment				
151	CORDIS CORPORATION	09/04/01	BPV	349862	Caval Thrombosis				
152	CORDIS CORPORATION	08/21/01	JWL	347962	Filter embolization	Migration (Rt Atrium)			
153	CORDIS CORPORATION	08/13/01	BPV	347020	Failed Deployment	Stuck in Sheath			
154	CORDIS CORPORATION	07/19/01	BPV	343267	Failed Deployment	Stuck in Sheath			
155	CORDIS CORPORATION	07/18/01	BPV	342518	Caval Thrombosis				
156	CORDIS CORPORATION	06/28/01	JWL	339597	Filter embolization	Caval thromb. less impt	Death	1	
157	CORDIS CORPORATION	06/28/01	BPV	339597	Caval thrombosis				
158	CORDIS CORPORATION	06/21/01	BPV	338397	Caval Thrombosis				
159	CORDIS CORPORATION	06/14/01	BPV	337233	Caval Thrombosis				
160	CORDIS CORPORATION	06/05/01	BPV	336244	Failed deployment				
161	CORDIS CORPORATION	05/24/01	BPV	334952	Partial Deployment				
162	CORDIS CORPORATION	05/21/01	BPV	333898	Caval Thrombosis				
163	CORDIS CORPORATION	05/21/01	BPV	334342	Partial Deployment				
164	CORDIS CORPORATION	05/18/01	BPV	333677	Caval Thrombosis				
165	CORDIS CORPORATION	05/01/01	BPV	329850	Caval Thrombosis				
166	CORDIS CORPORATION	04/27/01	BPV	329103	Caval Perforation	Sheath Perforation	Death		
167	CORDIS CORPORATION	04/27/01	BPV	329285	Caval Thrombosis				
168	CORDIS CORPORATION	04/27/01	BPV	329295	Partial Deployment				
169	CORDIS CORPORATION	04/12/01	BPV	326474	Caval Thrombosis				
170	CORDIS CORPORATION	04/12/01	BPV	326483	Caval Thrombosis				
171	CORDIS CORPORATION	04/12/01	BPV	326487	Caval Thrombosis				
172	CORDIS CORPORATION	04/12/01	BPV	326512	= Europe report				
173	CORDIS CORPORATION	04/02/01	BPV	323875	Caval Thrombosis	also had PE	Death		
174	CORDIS CORPORATION	03/27/01	BPV	322621	Caval Thrombosis		Death		
175	CORDIS CORPORATION	03/09/01	JWL	319700	Filter embolization	Migration (Rt Atrium)			
176	CORDIS CORPORATION	03/05/01	BPV	318548	Filter Fracture				
177	CORDIS CORPORATION	02/16/01	BPV	316469	Migration	Migration (just below renals)			
178	CORDIS CORPORATION	02/16/01	BPV	315469	Caval thrombosis	Migration (just below renals)			
179	CORDIS CORPORATION	02/08/01	BPV	315103	Failed deployment				
180	CORDIS CORPORATION	02/07/01	BPV	315120	Caval Thrombosis				
181	CORDIS CORPORATION	02/01/01	BPV	314825	Partial Deployment				
182	CORDIS CORPORATION	01/22/01	BPV	313140	Caval Perforation				
183	CORDIS CORPORATION	01/03/01	BPV	311191	Partial Deployment				
184	CORDIS CORPORATION	12/28/00	BPV	310050	Pulmonary Embolism				
185	CORDIS CORPORATION	12/20/00	BPV	309224	Pulmonary Embolism				
186	CORDIS CORPORATION	12/20/00	BPV	309224	Caval Thrombosis		Death		
187	CORDIS CORPORATION	12/14/00	BPV	308776	Caval Thrombosis				
188	CORDIS CORPORATION	12/07/00	BPV	307581	Caval Thrombosis				
189	CORDIS CORPORATION	12/04/00	BPV	307481	Partial Deployment				
190	CORDIS CORPORATION	11/27/00	BPV	306053	Caval Thrombosis				
191	CORDIS CORPORATION	11/27/00	BPV	306060	Caval Thrombosis				
192	CORDIS CORPORATION	11/27/00	BPV	306449	Caval Thrombosis				
193	CORDIS CORPORATION	11/13/00	BPV	303960	Caval Perforation				
194	CORDIS CORPORATION	11/10/00	BPV	304534	Caval Thrombosis				
195	CORDIS CORPORATION	11/02/00	BPV	303408	Partial Deployment				
196	CORDIS CORPORATION	10/24/00	BPV	301346	Caval Thrombosis				
197	CORDIS CORPORATION	09/25/00	BPV	297255	Partial Deployment				
198	CORDIS CORPORATION	09/15/00	BPV	296521	Pulmonary Embolism				
199									
200									
201									
202									
203									

[illegible]

EXHIBIT 6

**To Plaintiffs' Memorandum Re
Relevancy and Discoverability of FDA
Inspection and Warning Letter and
Recovery Cone Removal System**

(FILED UNDER SEAL)

EXHIBIT 7

**To Plaintiffs' Memorandum Re
Relevancy and Discoverability of FDA
Inspection and Warning Letter and
Recovery Cone Removal System**

(FILED UNDER SEAL)

EXHIBIT 8

**To Plaintiffs' Memorandum Re
Relevancy and Discoverability of FDA
Inspection and Warning Letter and
Recovery Cone Removal System**

(FILED UNDER SEAL)

EXHIBIT 9

**To Plaintiffs' Memorandum Re
Relevancy and Discoverability of FDA
Inspection and Warning Letter and
Recovery Cone Removal System**

(FILED UNDER SEAL)

EXHIBIT 10

**To Plaintiffs' Memorandum Re
Relevancy and Discoverability of FDA
Inspection and Warning Letter and
Recovery Cone Removal System**

(FILED UNDER SEAL)

EXHIBIT 11

**To Plaintiffs' Memorandum Re
Relevancy and Discoverability of FDA
Inspection and Warning Letter and
Recovery Cone Removal System**

1 IN THE UNITED STATES DISTRICT COURT

2 FOR THE DISTRICT OF ARIZONA

3

4

IN RE: BARD IVC FILTERS PRODUCTS)

5 LIABILITY LITIGATION) MD No.: 02641

_____)

6

7

8

9

10 DO NOT DISCLOSE - SUBJECT TO FURTHER CONFIDENTIALITY REVIEW

11

12

13 CONTINUED VIDEOTAPED DEPOSITION OF CHAD MICHAEL MODRA

14

15

Phoenix, Arizona

16 January 20, 2016

9:01 a.m.

17

18

19

20

21

22

23 REPORTED BY:

24 Robin L. B. Osterode, RPR, CSR

25 AZ Certified Reporter No. 50695

1 need -- you can take MAUDE completely out of the
2 picture. You don't even need MAUDE, if the -- for
3 the company to determine a rate comparison between
4 all of these complications, for all of its IVC
5 devices, in comparing those rates to the Simon
6 Nitinol Filter. Right?

7 A. I'm not --

8 Q. You don't understand the question?

9 A. No, I didn't under --

10 Q. You just told me you don't need MAUDE for
11 you to do internal rates of complications, because
12 you have actual sales. You have actual complaints.

13 A. Correct.

14 Q. All right. Take MAUDE. We don't need
15 MAUDE. MAUDE could burn. And that's true for the
16 Simon Nitinol Filter through and including the Denali
17 Filter?

18 A. That's true.

19 Q. You have all that data, sales data, you
20 have complaint data?

21 A. Yes, we have --

22 Q. And you can do internal rate comparison,
23 using your own data, that you know is more reliable
24 than what's in MAUDE. Right?

25 A. With -- yeah, with some caveats because we

EXHIBIT 12

**To Plaintiffs' Memorandum Re
Relevancy and Discoverability of FDA
Inspection and Warning Letter and
Recovery Cone Removal System**

IN THE UNITED STATES DISTRICT COURT
IN THE SOUTHERN DISTRICT OF FLORIDA

GOLKOW TECHNOLOGIES, INC.
877.370.3377 ph | 917.591.5672 fax
deps@golkow.com

Confidential - Subject to Protective Order

Page 165

1 Q. All right. Number two, the distraught customer
2 wants to know what system is in place to monitor
3 complication events. Is he or she entitled to have an
4 answer to that question?

5 A. I believe I already answered that. Yes. I think
6 that would be something that could be communicated.

7 Q. Number three, the distraught customer wants to
8 know what is the official complication rate of the G2
9 filter compared to other filters. Is he or she entitled
10 to have an answer to that question?

11 A. No. It's speculation.

12 Q. Why is it speculation?

13 A. Because I don't know the rates of other filters.

14 Q. Would Bard at least provide this distraught
15 customer, this physician, with its official complication
16 rate of the G2 as of August 2008?

17 A. No. It's confidential information.

18 Q. Okay. Number four, this distraught customer
19 wanted to know why are the fractures reported on Maude
20 similar. Is he or she entitled to have an answer to that
21 question?

22 MR. BORANIAN: I relodge my objections. It's
23 calling for speculation, improper form.

24 MR. JOHNSON: Your objection is preserved.

25 A. I don't understand the question, so I don't know